

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR042423\
 Data File : PR061005.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Apr 2023 01:40
 Operator : AJ\MA
 Sample : 02419-28
 Misc :
 ALS Vial : 57 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 25 03:09:01 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR040323CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 04 04:06:01 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.674	2.902	212.0E6	142.1E6	62.149	30.644 #
2)	SA Decachlor...	9.656	8.122	446.6E6	671.1E6	154.490	152.253
Target Compounds							
3)	L1 AR-1016-1	4.922	4.032	17538.7E6	53764.7E6	159984.534	350120.530 #
4)	L1 AR-1016-2	4.945	4.032	29834.5E6	53764.7E6	191434.523	241529.697 #
5)	L1 AR-1016-3	5.010	4.204	5565.3E6	24543.1E6	56529.768	213300.543 #
6)	L1 AR-1016-4	5.112	4.255	14900.1E6	25855.5E6	189027.918	287515.699 #
7)	L1 AR-1016-5	5.431	4.471	19178.8E6	30742.7E6	244843.322	261690.080
8)	L2 AR-1221-1	3.906	3.121	191.0E6	337.1E6	4607.421	6794.274 #
9)	L2 AR-1221-2	4.000	3.209	266.3E6	385.2E6	9321.064	10785.867
10)	L2 AR-1221-3	4.079	3.284	1415.1E6	2170.3E6	15273.885	17963.424
11)	L3 AR-1232-1	4.079	3.284	1415.1E6	2170.3E6	18200.109	21584.814
12)	L3 AR-1232-2	4.635	4.032	9057.2E6	53764.7E6	247341.868	553175.213 #
13)	L3 AR-1232-3	4.945	4.204	29834.5E6	24543.1E6	416432.326	498982.047
14)	L3 AR-1232-4	5.112	4.295	14900.1E6	25618.5E6	408525.802	595516.799 #
15)	L3 AR-1232-5	5.217	4.471	15724.2E6	30742.7E6	580138.538	632638.687
16)	L4 AR-1242-1	4.922	4.032	17538.7E6	53764.7E6	190702.684	431191.216 #
17)	L4 AR-1242-2	4.945	4.032	29834.5E6	53764.7E6	231703.851	299383.019 #
18)	L4 AR-1242-3	5.010	4.204	5565.3E6	24543.1E6	67466.979	264881.818 #
19)	L4 AR-1242-4	5.112	4.295	14900.1E6	25618.5E6	223115.078	291650.877 #
20)	L4 AR-1242-5	5.908	4.843	19426.3E6	34086.5E6	281148.731	293593.949
21)	L5 AR-1248-1	4.922	4.032	17538.7E6	53764.7E6	249414.324	558761.495 #
22)	L5 AR-1248-2	5.217	4.255	15724.2E6	25855.5E6	164941.908	192082.345
23)	L5 AR-1248-3	5.431	4.295	19178.8E6	25618.5E6	172335.259	188078.307
24)	L5 AR-1248-4	5.867	4.471	18426.8E6	30742.7E6	147080.009	184703.907 #
25)	L5 AR-1248-5	5.908	4.881	19426.3E6	27470.2E6	160266.829	161038.753
26)	L6 AR-1254-1	5.837	4.843	20377.7E6	34086.5E6	172403.469	134927.539
27)	L6 AR-1254-2	6.077	4.997	26747.2E6	26570.4E6	141079.621	121116.081
28)	L6 AR-1254-3	6.472	5.420	23288.9E6	36310.3E6	114975.744	99489.273
29)	L6 AR-1254-4	6.786	5.661	17876.9E6	28065.0E6	117379.244	122278.825
30)	L6 AR-1254-5	7.243	6.105	19284.6E6	38264.6E6	117934.146	115361.452
31)	L7 AR-1260-1	6.653	5.556	9373.6E6	26160.9E6	60691.555	102701.975 #
32)	L7 AR-1260-2	6.937	5.760	11172.6E6	18616.4E6	60618.606	59321.921
33)	L7 AR-1260-3	7.328	5.913	2402.8E6	18968.1E6	17602.961	65875.971 #
34)	L7 AR-1260-4	7.554	6.419	8540.3E6	4272.7E6	54027.638	17438.460 #
35)	L7 AR-1260-5	7.913	6.684	5974.1E6	10929.0E6	19535.426	18482.396
36)	L8 AR-1262-1	7.328	6.145	2402.8E6	4150.9E6	12017.349	11885.576
37)	L8 AR-1262-2	7.913	6.419	5974.1E6	4272.7E6	16797.496	13474.144
38)	L8 AR-1262-3	8.229	6.987	5816.6E6	4689.2E6	23637.052	18704.658
39)	L8 AR-1262-4	8.312	7.054	3248.1E6	11523.1E6	17340.537	23661.955 #
40)	L8 AR-1262-5	8.946	7.591	2240.8E6	3975.0E6	16468.886	17882.467
41)	L9 AR-1268-1	8.229	6.987	5816.6E6	4689.2E6	17322.687	7812.166 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR042423\
Data File : PR061005.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Apr 2023 01:40
Operator : AJ\MA
Sample : 02419-28
Misc :
ALS Vial : 57 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 25 03:09:01 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR040323CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 04 04:06:01 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42)	L9 AR-1268-2	8.312	7.054	3248.1E6	11523.1E6	10633.027	21085.988 #
43)	L9 AR-1268-3	8.538	7.275	892.2E6	640.2E6	3388.173	1381.812 #
44)	L9 AR-1268-4	8.946	7.591	2240.8E6	3975.0E6	18592.122	20331.883
45)	L9 AR-1268-5	9.340	7.884	2799.0E6	4669.8E6	3187.275	3076.433

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Instrument :
ECD_R
ClientSampleId :

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

