

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR042423\
Data File : PR060980.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Apr 2023 19:33
Operator : AJ\MA
Sample : 02418-20MS
Misc :
ALS Vial : 37 Sample Multiplier: 1


Instrument :
ECD_R
ClientSampleId :
EW8Z5MS

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 04/26/2023
Supervised By :Ankita Jodhani 04/26/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 25 02:52:15 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR040323CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 04 04:06:01 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(3) AR-1016-1 (L1)

R.T.	Response	Conc
4.92	48128262	439.02
4.94	69508833	446.01
5.01	45955006	466.79
5.11	42468881	538.77
5.43	37192968	474.82

(3) AR-1016-1 #2 (L1)

R.T.	Response	Conc
4.01	70957088	462.08
4.02	90760125	407.73
4.20	45879090	398.73
4.25	33776714	375.60
4.47	44477002	378.60

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR042423\
Data File : PR060980.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Apr 2023 19:33
Operator : AJ\MA
Sample : 02418-20MS
Misc :
ALS Vial : 37 Sample Multiplier: 1


Instrument :
ECD_R
ClientSampleId :
EW8Z5MS

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 04/26/2023
Supervised By :Ankita Jodhani 04/26/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 25 02:52:15 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR040323CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 04 04:06:01 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(3) AR-1016-1 #2 (L1)
R.T. Response Conc
4.92 48128262 439.02
4.94 67333709 432.05
5.01 41458189 421.12
5.11 36301896 460.54
5.43 37192968 474.82

(3) AR-1016-1 #2 (L1)
R.T. Response Conc
4.01 62631997 407.87
4.02 85425727 383.76
4.20 45879090 398.73
4.25 33776714 375.60
4.47 44477002 378.60

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR042423\
 Data File : PR060980.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Apr 2023 19:33
 Operator : AJ\MA
 Sample : 02418-20MS
 Misc :
 ALS Vial : 37 Sample Multiplier: 1



Instrument :
 ECD_R
 ClientSampleId :
 EW8Z5MS

Manual IntegrationsAPPROVED

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 25 02:52:15 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR040323CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 04 04:06:01 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Reviewed By :Yogesh Patel 04/26/2023
 Supervised By :Ankita Jodhani 04/26/2023

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(31) AR-1260-1 (L7)
 R.T. Response Conc
 6.65 66201073 428.64
 6.94 89232067 484.14
 7.33 57947266 424.52
 7.57 70502212 446.01
 7.92 121644430 397.78

(31) AR-1260-1 #2 (L7)
 R.T. Response Conc
 5.55 121857772 478.39
 5.76 145434404 463.43
 5.91 109180064 379.18
 6.42 92386708 377.07
 6.68 198525501 335.73

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR042423\
Data File : PR060980.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Apr 2023 19:33
Operator : AJ\MA
Sample : 02418-20MS
Misc :
ALS Vial : 37 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
EW8Z5MS

Manual IntegrationsAPPROVED

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 25 02:52:15 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR040323CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 04 04:06:01 2023
Response via : Initial Calibration
Integrator: ChemStation

Reviewed By :Yogesh Patel 04/26/2023
Supervised By :Ankita Jodhani 04/26/2023

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(31) AR-1260-1 #2 (L7)
R.T. Response Conc
6.65 66201073 428.64
6.94 89232067 484.14
7.33 57947266 424.52
7.57 80182333 507.25
7.92 121644430 397.78

(31) AR-1260-1 #2 (L7)
R.T. Response Conc
5.55 121857772 478.39
5.76 145434404 463.43
5.91 109180064 379.18
6.42 92386708 377.07
6.68 198525501 335.73