

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR042423\
 Data File : PR061005.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Apr 2023 01:40
 Operator : AJ\MA
 Sample : 02419-28
 Misc :
 ALS Vial : 57 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 EW925

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/26/2023
 Supervised By :Ankita Jodhani 04/26/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 25 03:09:01 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR040323CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 04 04:06:01 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

1) SA Tetrachlo...	3.683	2.902	74429856	142.1E6	21.822m	30.644 #
2) SA Decachlor...	9.656	8.122	446.6E6	671.1E6	154.490	152.253

Target Compounds

16) L4 AR-1242-1	4.922	4.003	17538.7E6	21452.1E6	190702.684	172045.274m
17) L4 AR-1242-2	4.945	4.035	29834.5E6	30549.2E6	231703.851	170109.727m#
18) L4 AR-1242-3	5.010	4.204	5565.3E6	24543.1E6	67466.979	264881.818 #
19) L4 AR-1242-4	5.112	4.295	14900.1E6	25618.5E6	223115.078	291650.877 #
20) L4 AR-1242-5	5.908	4.843	19426.3E6	34086.5E6	281148.731	293593.949
26) L6 AR-1254-1	5.837	4.843	20377.7E6	34086.5E6	172403.469	134927.539
27) L6 AR-1254-2	6.077	4.997	26747.2E6	26570.4E6	141079.621	121116.081
28) L6 AR-1254-3	6.472	5.420	23288.9E6	36310.3E6	114975.744	99489.273
29) L6 AR-1254-4	6.786	5.661	17876.9E6	28065.0E6	117379.244	122278.825
30) L6 AR-1254-5	7.243	6.105	19284.6E6	38264.6E6	117934.146	115361.452

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR042423\
Data File : PR061005.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Apr 2023 01:40
Operator : AJ\MA
Sample : 02419-28
Misc :
ALS Vial : 57 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
EW925

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/26/2023
Supervised By :Ankita Jodhani 04/26/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 25 03:09:01 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR040323CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 04 04:06:01 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

