

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR042424\
 Data File : PR066403.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Apr 2024 14:11
 Operator : AJ\MA
 Sample : AR1248ICC050
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR1248ICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 24 15:06:13 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR042424.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 24 14:52:25 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.660	2.897	23944806	20999628	4.717	4.725
2) SA Decachlor...	9.559	8.131	10690570	22973227	4.945	5.007
Target Compounds						
21) L5 AR-1248-1	4.880	4.008	4358803	4456475	50.363	47.934
22) L5 AR-1248-2	5.171	4.254	6885635	6782371	51.731	50.122
23) L5 AR-1248-3	5.382	4.295	7589956	6840092	50.565	49.679
24) L5 AR-1248-4	5.815	4.470	7038924	8182871	49.402	49.503
25) L5 AR-1248-5	5.854	4.880	7092661	8050511	49.526	49.889



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR042424\
Data File : PR066403.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Apr 2024 14:11
Operator : AJ\MA
Sample : AR1248ICC050
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR1248ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 24 15:06:13 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR042424.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 24 14:52:25 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.660 min
Delta R.T.: 0.000 min
Response: 23944806
Conc: 4.72 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1248ICC050




#1 Tetrachloro-m-xylene

R.T.: 2.897 min
Delta R.T.: 0.000 min
Response: 20999628
Conc: 4.73 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.559 min
Delta R.T.: 0.000 min
Response: 10690570
Conc: 4.95 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.131 min
Delta R.T.: 0.000 min
Response: 22973227
Conc: 5.01 ng/ml



#21 AR-1248-1


R.T.: 4.880 min
Delta R.T.: 0.000 min
Response: 4358803
Conc: 50.36 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1248ICC050




#21 AR-1248-1

R.T.: 4.008 min
Delta R.T.: 0.000 min
Response: 4456475
Conc: 47.93 ng/ml




#22 AR-1248-2

R.T.: 5.171 min
Delta R.T.: 0.000 min
Response: 6885635
Conc: 51.73 ng/ml



#22 AR-1248-2


R.T.: 4.254 min
Delta R.T.: 0.000 min
Response: 6782371
Conc: 50.12 ng/ml



#23 AR-1248-3


R.T.: 5.382 min
Delta R.T.: 0.000 min
Response: 7589956
Conc: 50.57 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1248ICC050




#23 AR-1248-3

R.T.: 4.295 min
Delta R.T.: 0.000 min
Response: 6840092
Conc: 49.68 ng/ml




#24 AR-1248-4

R.T.: 5.815 min
Delta R.T.: 0.000 min
Response: 7038924
Conc: 49.40 ng/ml



#24 AR-1248-4


R.T.: 4.470 min
Delta R.T.: 0.000 min
Response: 8182871
Conc: 49.50 ng/ml



#25 AR-1248-5

R.T.: 5.854 min
Delta R.T.: 0.000 min
Response: 7092661
Conc: 49.53 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1248ICC050



#25 AR-1248-5

R.T.: 4.880 min
Delta R.T.: 0.000 min
Response: 8050511
Conc: 49.89 ng/ml