

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR042424\
 Data File : PR066397.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Apr 2024 12:42
 Operator : AJ\MA
 Sample : AR1242ICC250
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR1242ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 24 14:43:19 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR042424.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 24 14:31:37 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.662	2.897	133.1E6	115.4E6	25.376	25.657
2)	SA Decachlor...	9.571	8.132	60451491	122.7E6	25.863	26.323
Target Compounds							
16)	L4 AR-1242-1	4.881	4.007	30989993	33386986	265.084	266.304
17)	L4 AR-1242-2	4.904	4.025	48090003	47126587	267.452	264.599
18)	L4 AR-1242-3	4.967	4.203	31889579	25997592	269.584	266.720
19)	L4 AR-1242-4	5.069	4.295	24711332	26105426	263.848	273.143
20)	L4 AR-1242-5	5.855	4.838	23863194	32900244	269.434	268.644



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR042424\
Data File : PR066397.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Apr 2024 12:42
Operator : AJ\MA
Sample : AR1242ICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR1242ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 24 14:43:19 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR042424.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 24 14:31:37 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.662 min
Delta R.T.: 0.000 min
Response: 133093355
Conc: 25.38 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1242ICC250




#1 Tetrachloro-m-xylene

R.T.: 2.897 min
Delta R.T.: 0.000 min
Response: 115448964
Conc: 25.66 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.571 min
Delta R.T.: 0.000 min
Response: 60451491
Conc: 25.86 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.132 min
Delta R.T.: 0.000 min
Response: 122661992
Conc: 26.32 ng/ml



#16 AR-1242-1


R.T.: 4.881 min
Delta R.T.: 0.000 min
Response: 30989993
Conc: 265.08 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1242ICC250




#16 AR-1242-1

R.T.: 4.007 min
Delta R.T.: 0.000 min
Response: 33386986
Conc: 266.30 ng/ml




#17 AR-1242-2

R.T.: 4.904 min
Delta R.T.: 0.000 min
Response: 48090003
Conc: 267.45 ng/ml



#17 AR-1242-2


R.T.: 4.025 min
Delta R.T.: 0.000 min
Response: 47126587
Conc: 264.60 ng/ml



#18 AR-1242-3


R.T.: 4.967 min
Delta R.T.: 0.000 min
Response: 31889579
Conc: 269.58 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1242ICC250




#18 AR-1242-3

R.T.: 4.203 min
Delta R.T.: 0.000 min
Response: 25997592
Conc: 266.72 ng/ml




#19 AR-1242-4

R.T.: 5.069 min
Delta R.T.: 0.000 min
Response: 24711332
Conc: 263.85 ng/ml



#19 AR-1242-4


R.T.: 4.295 min
Delta R.T.: 0.000 min
Response: 26105426
Conc: 273.14 ng/ml



#20 AR-1242-5

R.T.: 5.855 min
Delta R.T.: 0.000 min
Response: 23863194
Conc: 269.43 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1242ICC250



#20 AR-1242-5

R.T.: 4.838 min
Delta R.T.: 0.000 min
Response: 32900244
Conc: 268.64 ng/ml