

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR042723\
 Data File : PR061126.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Apr 2023 10:38
 Operator : AJ\MA
 Sample : 02419-24DL 20X
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 EW921DL

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/01/2023
 Supervised By :Ankita Jodhani 05/01/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 27 18:42:00 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR040323CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 04 04:06:01 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.684	2.900	4637055	6307052	1.360m	1.360
2) SA Decachlor...	9.667	8.123	11089857	29002360	3.836	6.580 #
Target Compounds						
16) L4 AR-1242-1	4.924	4.007	582.8E6	852.4E6	6336.998	6836.464
17) L4 AR-1242-2	4.947	4.024	1000.4E6	1481.8E6	7769.756	8251.157
18) L4 AR-1242-3	5.012	4.201	186.9E6	668.8E6	2266.017	7218.119 #
19) L4 AR-1242-4	5.115	4.293	451.4E6	675.5E6	6759.276	7689.593
20) L4 AR-1242-5	5.910	4.836	562.4E6	1356.5E6	8138.718	11683.978 #
26) L6 AR-1254-1	5.839	4.836	557.6E6	1356.5E6	4717.326	5369.628
27) L6 AR-1254-2	6.080	4.994	790.9E6	626.2E6	4171.640	2854.330 #
28) L6 AR-1254-3	6.475	5.412	603.6E6	1073.9E6	2980.180	2942.395
29) L6 AR-1254-4	6.790	5.658	443.1E6	697.9E6	2909.664	3040.831
30) L6 AR-1254-5	7.248	6.101	421.9E6	841.9E6	2580.324	2538.280

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR042723\
Data File : PR061126.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Apr 2023 10:38
Operator : AJ\MA
Sample : 02419-24DL 20X
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
EW921DL

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/01/2023
Supervised By :Ankita Jodhani 05/01/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 27 18:42:00 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR040323CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 04 04:06:01 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

