

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR042723\
 Data File : PR061128.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Apr 2023 11:36
 Operator : AJ\MA
 Sample : 02447-15DL 20X
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 EW935DL

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/01/2023
 Supervised By :Ankita Jodhani 05/01/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 27 18:42:22 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR040323CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 04 04:06:01 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.681	2.900	4539123	5877745	1.331m	1.268
2)	SA Decachlor...	9.659	8.123	8842900	11782259	3.059	2.673
Target Compounds							
16)	L4 AR-1242-1	4.920	4.007	508.2E6	748.4E6	5525.668	6002.099
17)	L4 AR-1242-2	4.943	4.024	883.5E6	1295.5E6	6861.816	7213.587
18)	L4 AR-1242-3	5.008	4.201	203.0E6	613.9E6	2460.638	6625.788 #
19)	L4 AR-1242-4	5.111	4.293	403.8E6	601.8E6	6046.828	6850.542
20)	L4 AR-1242-5	5.907	4.836	514.8E6	1336.5E6	7451.173	11511.956 #
26)	L6 AR-1254-1	5.836	4.836	564.0E6	1336.5E6	4771.410	5290.572
27)	L6 AR-1254-2	6.076	4.994	798.9E6	689.2E6	4213.593	3141.490 #
28)	L6 AR-1254-3	6.471	5.412	653.6E6	1171.6E6	3226.698	3210.274
29)	L6 AR-1254-4	6.785	5.658	477.3E6	761.5E6	3133.758	3317.787
30)	L6 AR-1254-5	7.244	6.101	455.6E6	981.7E6	2786.282	2959.537

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR042723\
Data File : PR061128.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Apr 2023 11:36
Operator : AJ\MA
Sample : 02447-15DL 20X
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
EW935DL

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/01/2023
Supervised By :Ankita Jodhani 05/01/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 27 18:42:22 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR040323CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 04 04:06:01 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

