

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR050223\
 Data File : PR061154.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 May 2023 21:04
 Operator : YP\AJ
 Sample : AR1242ICC200
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR12422202

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 03 02:46:18 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR050223CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed May 03 02:45:26 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.684	2.897	39133210	50227354	10.373	10.016
2) SA Decachlor...	9.658	8.123	59718916	95415191	21.518	20.436
Target Compounds						
16) L4 AR-1242-1	4.918	4.005	22302418	27433804	220.517	202.183
17) L4 AR-1242-2	4.941	4.022	29042730	39015893	207.261	205.132
18) L4 AR-1242-3	5.006	4.199	18894508	20662104	209.676	206.607
19) L4 AR-1242-4	5.109	4.292	15201320	19896069	208.814	211.960
20) L4 AR-1242-5	5.904	4.834	16497002	25673630	213.204	206.593

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR050223\
Data File : PR061154.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 May 2023 21:04
Operator : YP\AJ
Sample : AR1242ICC200
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR12422202

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 03 02:46:18 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR050223CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed May 03 02:45:26 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

