

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050323\
 Data File : PD075203.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 May 2023 08:29
 Operator : AR\AJ
 Sample : 02447-19MSD (Sig #1); 02447-21MSD (Sig #2)
 Misc :
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 04 10:59:20 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
 Quant Title : GC Extractables
 QLast Update : Thu May 04 02:14:07 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.551	2.772	27227042	18754598	11.937	17.781 #
27)	SA Decachlor...	9.082	7.917	794.4E6	95384788	320.989	101.002 #
Target Compounds							
2)	A alpha-BHC	4.007	3.274	147.2E6	60704149	38.272	39.189
3)	MA gamma-BHC...	4.314	3.609	569.9E6	455.1E6	158.512	309.605 #
4)	MA Heptachlor	4.932	3.945	-60923736	207.4E6	N.D.	141.952
5)	MB Aldrin	5.274	4.227	167.2E6	81950733	45.839	55.257
6)	B beta-BHC	4.536	3.901	41248911	18379790	25.129	25.242
7)	B delta-BHC	4.787	4.132	104.3E6	117.2E6	30.869	74.149 #
8)	B Heptachlo...	5.699	4.728	277.4E6	38737943	81.858	27.774 #
9)	A Endosulfan I	6.090	5.128	82245385	612.7E6	27.266	510.539 #
10)	B trans-Chl...	5.958	4.980	-17420394	107.6E6	N.D.	78.852
11)	B cis-Chlor...	6.033	5.045	667.9E6	105.9E6	195.980	75.901 #
12)	B 4,4'-DDE	6.211	5.234	592.5E6	177.9E6	181.442	138.382
13)	MA Dieldrin	6.367	5.369	555.3E6	323.7E6	162.376	237.245 #
14)	MA Endrin	6.598	5.640	618.9E6	190.2E6	214.025	156.382 #
15)	B Endosulfa...	6.814	5.936	227.3E6	212.0E6	70.574	165.709 #
16)	A 4,4'-DDD	6.723	5.788	718.9E6	128.8E6	297.489	129.939 #
17)	MA 4,4'-DDT	7.045	6.039	1266.9E6	372.5E6	499.139	369.594 #
18)	B Endrin al...	6.945	6.114	283.4E6	-18214994	119.806	N.D. #
19)	B Endosulfa...	7.180	6.340	185.5E6	122.1E6	67.450	104.985 #
20)	A Methoxychlor	7.519	6.617	650.8E6	191.9E6	443.026	335.116
21)	B Endrin ke...	7.666	6.843	130.2E6	78177424	42.940	59.288 #
22)	Toxaphene-1	6.269	5.009	147.2E6	63031768	6396.842	9126.141 #
23)	Toxaphene-2	6.479	5.730f	19318836	232.6E6	647.435	29135.761 #
24)	Toxaphene-3	6.723f	5.936	718.9E6	212.0E6	34080.835	24199.503 #
25)	Toxaphene-4	7.180	6.617	185.5E6	191.9E6	2846.033	7581.602 #
26)	Toxaphene-5	7.633	6.689f	118.6E6	8534128	2500.096	234.983 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD050323\
Data File : PD075203.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 May 2023 08:29
Operator : AR\AJ
Sample : 02447-19MSD (Sig #1); 02447-21MSD (Sig #2)
Misc :
ALS Vial : 41 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 04 10:59:20 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD050323CLP.M
Quant Title : GC Extractables
QLast Update : Thu May 04 02:14:07 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

