

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR050618\
 Data File : PR027733.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 May 2018 11:19
 Operator : UA/SM
 Sample : J2133-09
 Misc : M8011-504.1 MDL; 0.025 PPB
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 MDL-WATER-03-QT2-2018

Manual Integrations
 APPROVED

ugo
 6/8/2018 7:28:17 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 11 12:57:34 2018
 Quant Method : P:\HPCHEM1\ECD_R\Method\PR050618-504.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sun May 06 19:21:29 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

Target Compounds						
1) SA EDB	3.260	2.533	1337083	1852540	0.021m	0.020m
2) SA DBCP	7.327	5.774	3166389	6833047	0.027m	0.024m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR050618\
Data File : PR027733.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 May 2018 11:19
Operator : UA/SM
Sample : J2133-09
Misc : M8011-504.1 MDL; 0.025 PPB
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
MDL-WATER-03-QT2-2018

Manual Integrations
APPROVED

ugo
6/8/2018 7:28:17 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 11 12:57:34 2018
Quant Method : P:\HPCHEM1\ECD_R\Method\PR050618-504.M
Quant Title : GC EXTRACTABLES
QLast Update : Sun May 06 19:21:29 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

