

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR050724\
 Data File : PR066560.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 May 2024 13:27
 Operator : AJ\MA
 Sample : P2377-14
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 PCB-GCP2-BLANK-SPIKE

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 07 22:17:56 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR042324CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 24 04:40:00 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.683	2.873	248836	39355960	0.043	7.739 #
2)	SA Decachlor...	9.535	8.126	1745438	1837610	1.251	0.710 #
Target Compounds							
3)	L1 AR-1016-1	4.874	4.005	71455792	68430143	450.232	395.401
4)	L1 AR-1016-2	4.895	4.023	110.8E6	101.2E6	451.879	412.569
5)	L1 AR-1016-3	4.960	4.200	70879028	53143681	436.008	387.384
6)	L1 AR-1016-4	5.062	4.251	55583509	42118721	432.242	374.600
7)	L1 AR-1016-5	5.376	4.466	52875630	54195945	410.365	375.449
8)	L2 AR-1221-1	3.873	3.120	7987470	7408194	127.765	127.184
9)	L2 AR-1221-2	3.965	3.207	7282231	6674676	165.756	158.632
10)	L2 AR-1221-3	4.043	3.282	38925580	33500225	276.543	245.191
11)	L3 AR-1232-1	4.043	3.282	38925580	33500225	332.009	294.352
12)	L3 AR-1232-2	4.589	4.023	57651082	101.2E6	871.297	947.088
13)	L3 AR-1232-3	4.895	4.200	110.8E6	53143681	1010.880	914.096
14)	L3 AR-1232-4	5.062	4.292	55583509	50355865	962.447	955.798
15)	L3 AR-1232-5	5.165	4.466	44851509	54195945	1047.431	949.148
16)	L4 AR-1242-1	4.895	4.005	110.8E6	68430143	899.949	529.165 #
17)	L4 AR-1242-2	4.895	4.023	110.8E6	101.2E6	589.022	553.745
18)	L4 AR-1242-3	4.960	4.200	70879028	53143681	566.554	519.054
19)	L4 AR-1242-4	5.062	4.292	55583509	50355865	557.212	497.319
20)	L4 AR-1242-5	5.848	4.834	7531933	43145728	78.953	349.271 #
21)	L5 AR-1248-1	4.895	4.005	110.8E6	68430143	1206.652	691.588 #
22)	L5 AR-1248-2	5.165	4.251	44851509	42118721	317.073	287.961
23)	L5 AR-1248-3	5.376	4.292	52875630	50355865	329.090	338.695
24)	L5 AR-1248-4	5.806	4.466	6923646	54195945	44.734	306.297 #
25)	L5 AR-1248-5	5.848	4.876	7531933	6381607	46.568	37.643
26)	L6 AR-1254-1	5.778	4.834	40287715	43145728	224.772	161.724 #
27)	L6 AR-1254-2	6.014	4.992	43848325	41903592	166.482	177.876
28)	L6 AR-1254-3	6.404	5.431	21893864	77771195	83.710	208.798 #
29)	L6 AR-1254-4	6.713	5.657	13048042	8656868	76.762	38.492 #
30)	L6 AR-1254-5	7.200	6.100	16087749	169.2E6	85.034	488.562 #
31)	L7 AR-1260-1	6.584	5.551	99877359	116.0E6	429.165	401.355
32)	L7 AR-1260-2	6.865	5.756	106.9E6	138.5E6	429.331	401.999
33)	L7 AR-1260-3	7.254	5.913	66361793	132.5E6	348.199	416.889
34)	L7 AR-1260-4	7.495	6.418	76218762	99729354	372.172	358.817
35)	L7 AR-1260-5	7.834	6.682	130.2E6	237.3E6	384.596	381.889
36)	L8 AR-1262-1	7.254	6.144	66361793	100.9E6	272.208	275.631
37)	L8 AR-1262-2	7.834	6.682	130.2E6	237.3E6	393.320	375.482
38)	L8 AR-1262-3	8.145	6.986	73485186	41278881	336.104	154.217 #
39)	L8 AR-1262-4	8.224	7.052	17633577	171.9E6	104.010	356.450 #
40)	L8 AR-1262-5	8.845	7.590	24701939	51567097	246.232	225.422
41)	L9 AR-1268-1	8.145	6.986	73485186	41278881	190.905	54.284 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR050724\
Data File : PR066560.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 May 2024 13:27
Operator : AJ\MA
Sample : P2377-14
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
PCB-GCP2-BLANK-SPIKE

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 07 22:17:56 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR042324CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 24 04:40:00 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42)	L9 AR-1268-2	8.224	7.052	17633577	171.9E6	51.230	251.960 #
43)	L9 AR-1268-3	8.443	7.276	2544968	3869371	8.746	6.370 #
44)	L9 AR-1268-4	8.845	7.590	24701939	51567097	219.955	203.072
45)	L9 AR-1268-5	9.228	7.884	8221678	15879797	10.585	8.993



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR050724\
Data File : PR066560.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 May 2024 13:27
Operator : AJ\MA
Sample : P2377-14
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
PCB-GCP2-BLANK-SPIKE

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 07 22:17:56 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR042324CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 24 04:40:00 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#3 AR-1016-1


R.T.: 4.874 min
Delta R.T.: -0.010 min
Response: 71455792
Conc: 450.23 ng/ml

Instrument :
ECD_R
ClientSampleId :
PCB-GCP2-BLANK-SPIKE




#3 AR-1016-1

R.T.: 4.005 min
Delta R.T.: -0.003 min
Response: 68430143
Conc: 395.40 ng/ml




#4 AR-1016-2

R.T.: 4.895 min
Delta R.T.: -0.010 min
Response: 110766153
Conc: 451.88 ng/ml



#4 AR-1016-2


R.T.: 4.023 min
Delta R.T.: -0.003 min
Response: 101245484
Conc: 412.57 ng/ml



#5 AR-1016-3


R.T.: 4.960 min
Delta R.T.: -0.010 min
Response: 70879028
Conc: 436.01 ng/ml

Instrument :
ECD_R
ClientSampleId :
PCB-GCP2-BLANK-SPIKE




#5 AR-1016-3

R.T.: 4.200 min
Delta R.T.: -0.003 min
Response: 53143681
Conc: 387.38 ng/ml




#6 AR-1016-4

R.T.: 5.062 min
Delta R.T.: -0.010 min
Response: 55583509
Conc: 432.24 ng/ml



#6 AR-1016-4


R.T.: 4.251 min
Delta R.T.: -0.003 min
Response: 42118721
Conc: 374.60 ng/ml



#7 AR-1016-5


R.T.: 5.376 min
Delta R.T.: -0.010 min
Response: 52875630
Conc: 410.36 ng/ml

Instrument :
ECD_R
ClientSampleId :
PCB-GCP2-BLANK-SPIKE




#7 AR-1016-5

R.T.: 4.466 min
Delta R.T.: -0.004 min
Response: 54195945
Conc: 375.45 ng/ml




#8 AR-1221-1

R.T.: 3.873 min
Delta R.T.: -0.013 min
Response: 7987470
Conc: 127.76 ng/ml



#8 AR-1221-1


R.T.: 3.120 min
Delta R.T.: -0.001 min
Response: 7408194
Conc: 127.18 ng/ml



#9 AR-1221-2


R.T.: 3.965 min
Delta R.T.: -0.010 min
Response: 7282231
Conc: 165.76 ng/ml

Instrument :
ECD_R
ClientSampleId :
PCB-GCP2-BLANK-SPIKE




#9 AR-1221-2

R.T.: 3.207 min
Delta R.T.: 0.000 min
Response: 6674676
Conc: 158.63 ng/ml




#10 AR-1221-3

R.T.: 4.043 min
Delta R.T.: -0.010 min
Response: 38925580
Conc: 276.54 ng/ml



#10 AR-1221-3


R.T.: 3.282 min
Delta R.T.: 0.000 min
Response: 33500225
Conc: 245.19 ng/ml



#11 AR-1232-1


R.T.: 4.043 min
Delta R.T.: -0.010 min
Response: 38925580
Conc: 332.01 ng/ml

Instrument :
ECD_R
ClientSampleId :
PCB-GCP2-BLANK-SPIKE




#11 AR-1232-1

R.T.: 3.282 min
Delta R.T.: -0.001 min
Response: 33500225
Conc: 294.35 ng/ml




#12 AR-1232-2

R.T.: 4.589 min
Delta R.T.: -0.010 min
Response: 57651082
Conc: 871.30 ng/ml



#12 AR-1232-2


R.T.: 4.023 min
Delta R.T.: -0.003 min
Response: 101245484
Conc: 947.09 ng/ml



#13 AR-1232-3


R.T.: 4.895 min
Delta R.T.: -0.012 min
Response: 110766153
Conc: 1010.88 ng/ml

Instrument :
ECD_R
ClientSampleId :
PCB-GCP2-BLANK-SPIKE




#13 AR-1232-3

R.T.: 4.200 min
Delta R.T.: -0.003 min
Response: 53143681
Conc: 914.10 ng/ml




#14 AR-1232-4

R.T.: 5.062 min
Delta R.T.: -0.011 min
Response: 55583509
Conc: 962.45 ng/ml



#14 AR-1232-4


R.T.: 4.292 min
Delta R.T.: -0.004 min
Response: 50355865
Conc: 955.80 ng/ml



#15 AR-1232-5


R.T.: 5.165 min
Delta R.T.: -0.011 min
Response: 44851509
Conc: 1047.43 ng/ml

Instrument :
ECD_R
ClientSampleId :
PCB-GCP2-BLANK-SPIKE




#15 AR-1232-5

R.T.: 4.466 min
Delta R.T.: -0.004 min
Response: 54195945
Conc: 949.15 ng/ml




#16 AR-1242-1

R.T.: 4.895 min
Delta R.T.: 0.010 min
Response: 110766153
Conc: 899.95 ng/ml



#16 AR-1242-1


R.T.: 4.005 min
Delta R.T.: -0.003 min
Response: 68430143
Conc: 529.17 ng/ml



#17 AR-1242-2


R.T.: 4.895 min
Delta R.T.: -0.012 min
Response: 110766153
Conc: 589.02 ng/ml

Instrument :
ECD_R
ClientSampleId :
PCB-GCP2-BLANK-SPIKE




#17 AR-1242-2

R.T.: 4.023 min
Delta R.T.: -0.003 min
Response: 101245484
Conc: 553.74 ng/ml




#18 AR-1242-3

R.T.: 4.960 min
Delta R.T.: -0.011 min
Response: 70879028
Conc: 566.55 ng/ml



#18 AR-1242-3


R.T.: 4.200 min
Delta R.T.: -0.004 min
Response: 53143681
Conc: 519.05 ng/ml



#19 AR-1242-4


R.T.: 5.062 min
Delta R.T.: -0.011 min
Response: 55583509
Conc: 557.21 ng/ml

Instrument :
ECD_R
ClientSampleId :
PCB-GCP2-BLANK-SPIKE




#19 AR-1242-4

R.T.: 4.292 min
Delta R.T.: -0.004 min
Response: 50355865
Conc: 497.32 ng/ml




#20 AR-1242-5

R.T.: 5.848 min
Delta R.T.: -0.012 min
Response: 7531933
Conc: 78.95 ng/ml



#20 AR-1242-5


R.T.: 4.834 min
Delta R.T.: -0.006 min
Response: 43145728
Conc: 349.27 ng/ml



#21 AR-1248-1


R.T.: 4.895 min
Delta R.T.: 0.010 min
Response: 110766153
Conc: 1206.65 ng/ml

Instrument :
ECD_R
ClientSampleId :
PCB-GCP2-BLANK-SPIKE




#21 AR-1248-1

R.T.: 4.005 min
Delta R.T.: -0.003 min
Response: 68430143
Conc: 691.59 ng/ml




#22 AR-1248-2

R.T.: 5.165 min
Delta R.T.: -0.012 min
Response: 44851509
Conc: 317.07 ng/ml



#22 AR-1248-2


R.T.: 4.251 min
Delta R.T.: -0.004 min
Response: 42118721
Conc: 287.96 ng/ml



#23 AR-1248-3


R.T.: 5.376 min
Delta R.T.: -0.011 min
Response: 52875630
Conc: 329.09 ng/ml

Instrument :
ECD_R
ClientSampleId :
PCB-GCP2-BLANK-SPIKE




#23 AR-1248-3

R.T.: 4.292 min
Delta R.T.: -0.004 min
Response: 50355865
Conc: 338.69 ng/ml




#24 AR-1248-4

R.T.: 5.806 min
Delta R.T.: -0.014 min
Response: 6923646
Conc: 44.73 ng/ml



#24 AR-1248-4


R.T.: 4.466 min
Delta R.T.: -0.004 min
Response: 54195945
Conc: 306.30 ng/ml



#25 AR-1248-5


R.T.: 5.848 min
Delta R.T.: -0.012 min
Response: 7531933
Conc: 46.57 ng/ml

Instrument :
ECD_R
ClientSampleId :
PCB-GCP2-BLANK-SPIKE




#25 AR-1248-5

R.T.: 4.876 min
Delta R.T.: -0.005 min
Response: 6381607
Conc: 37.64 ng/ml




#26 AR-1254-1

R.T.: 5.778 min
Delta R.T.: -0.011 min
Response: 40287715
Conc: 224.77 ng/ml



#26 AR-1254-1


R.T.: 4.834 min
Delta R.T.: -0.005 min
Response: 43145728
Conc: 161.72 ng/ml



#27 AR-1254-2


R.T.: 6.014 min
Delta R.T.: -0.011 min
Response: 43848325
Conc: 166.48 ng/ml

Instrument :
ECD_R
ClientSampleId :
PCB-GCP2-BLANK-SPIKE




#27 AR-1254-2

R.T.: 4.992 min
Delta R.T.: -0.005 min
Response: 41903592
Conc: 177.88 ng/ml




#28 AR-1254-3

R.T.: 6.404 min
Delta R.T.: -0.013 min
Response: 21893864
Conc: 83.71 ng/ml



#28 AR-1254-3


R.T.: 5.431 min
Delta R.T.: 0.014 min
Response: 77771195
Conc: 208.80 ng/ml



#29 AR-1254-4


R.T.: 6.713 min
Delta R.T.: -0.015 min
Response: 13048042
Conc: 76.76 ng/ml

Instrument :
ECD_R
ClientSampleId :
PCB-GCP2-BLANK-SPIKE




#29 AR-1254-4

R.T.: 5.657 min
Delta R.T.: -0.006 min
Response: 8656868
Conc: 38.49 ng/ml




#30 AR-1254-5

R.T.: 7.200 min
Delta R.T.: 0.016 min
Response: 16087749
Conc: 85.03 ng/ml



#30 AR-1254-5


R.T.: 6.100 min
Delta R.T.: -0.007 min
Response: 169152554
Conc: 488.56 ng/ml



#31 AR-1260-1


R.T.: 6.584 min
Delta R.T.: -0.011 min
Response: 99877359
Conc: 429.16 ng/ml

Instrument :
ECD_R
ClientSampleId :
PCB-GCP2-BLANK-SPIKE




#31 AR-1260-1

R.T.: 5.551 min
Delta R.T.: -0.005 min
Response: 115980273
Conc: 401.36 ng/ml




#32 AR-1260-2

R.T.: 6.865 min
Delta R.T.: -0.011 min
Response: 106936616
Conc: 429.33 ng/ml



#32 AR-1260-2


R.T.: 5.756 min
Delta R.T.: -0.005 min
Response: 138506283
Conc: 402.00 ng/ml



#33 AR-1260-3


R.T.: 7.254 min
Delta R.T.: -0.013 min
Response: 66361793
Conc: 348.20 ng/ml

Instrument :
ECD_R
ClientSampleId :
PCB-GCP2-BLANK-SPIKE




#33 AR-1260-3

R.T.: 5.913 min
Delta R.T.: -0.006 min
Response: 132525637
Conc: 416.89 ng/ml




#34 AR-1260-4

R.T.: 7.495 min
Delta R.T.: -0.014 min
Response: 76218762
Conc: 372.17 ng/ml



#34 AR-1260-4


R.T.: 6.418 min
Delta R.T.: -0.005 min
Response: 99729354
Conc: 358.82 ng/ml



#35 AR-1260-5


R.T.: 7.834 min
Delta R.T.: -0.015 min
Response: 130247704
Conc: 384.60 ng/ml

Instrument :
ECD_R
ClientSampleId :
PCB-GCP2-BLANK-SPIKE




#35 AR-1260-5

R.T.: 6.682 min
Delta R.T.: -0.005 min
Response: 237283873
Conc: 381.89 ng/ml




#36 AR-1262-1

R.T.: 7.254 min
Delta R.T.: -0.008 min
Response: 66361793
Conc: 272.21 ng/ml



#36 AR-1262-1


R.T.: 6.144 min
Delta R.T.: -0.006 min
Response: 100898114
Conc: 275.63 ng/ml



#37 AR-1262-2


R.T.: 7.834 min
Delta R.T.: -0.009 min
Response: 130247704
Conc: 393.32 ng/ml

Instrument :
ECD_R
ClientSampleId :
PCB-GCP2-BLANK-SPIKE




#37 AR-1262-2

R.T.: 6.682 min
Delta R.T.: -0.006 min
Response: 237283873
Conc: 375.48 ng/ml




#38 AR-1262-3

R.T.: 8.145 min
Delta R.T.: -0.006 min
Response: 73485186
Conc: 336.10 ng/ml



#38 AR-1262-3


R.T.: 6.986 min
Delta R.T.: -0.006 min
Response: 41278881
Conc: 154.22 ng/ml



#39 AR-1262-4


R.T.: 8.224 min
Delta R.T.: -0.014 min
Response: 17633577
Conc: 104.01 ng/ml

Instrument :
ECD_R
ClientSampleId :
PCB-GCP2-BLANK-SPIKE




#39 AR-1262-4

R.T.: 7.052 min
Delta R.T.: -0.008 min
Response: 171865295
Conc: 356.45 ng/ml




#40 AR-1262-5

R.T.: 8.845 min
Delta R.T.: -0.017 min
Response: 24701939
Conc: 246.23 ng/ml



#40 AR-1262-5


R.T.: 7.590 min
Delta R.T.: -0.006 min
Response: 51567097
Conc: 225.42 ng/ml



#41 AR-1268-1


R.T.: 8.145 min
Delta R.T.: -0.002 min
Response: 73485186
Conc: 190.91 ng/ml

Instrument :
ECD_R
ClientSampleId :
PCB-GCP2-BLANK-SPIKE




#41 AR-1268-1

R.T.: 6.986 min
Delta R.T.: -0.005 min
Response: 41278881
Conc: 54.28 ng/ml




#42 AR-1268-2

R.T.: 8.224 min
Delta R.T.: -0.013 min
Response: 17633577
Conc: 51.23 ng/ml



#42 AR-1268-2


R.T.: 7.052 min
Delta R.T.: -0.008 min
Response: 171865295
Conc: 251.96 ng/ml



#43 AR-1268-3


R.T.: 8.443 min
Delta R.T.: -0.013 min
Response: 2544968
Conc: 8.75 ng/ml

Instrument :
ECD_R
ClientSampleId :
PCB-GCP2-BLANK-SPIKE




#43 AR-1268-3

R.T.: 7.276 min
Delta R.T.: -0.005 min
Response: 3869371
Conc: 6.37 ng/ml




#44 AR-1268-4

R.T.: 8.845 min
Delta R.T.: -0.015 min
Response: 24701939
Conc: 219.96 ng/ml



#44 AR-1268-4


R.T.: 7.590 min
Delta R.T.: -0.004 min
Response: 51567097
Conc: 203.07 ng/ml



#45 AR-1268-5

R.T.: 9.228 min
Delta R.T.: -0.017 min
Response: 8221678
Conc: 10.58 ng/ml

Instrument :
ECD_R
ClientSampleId :
PCB-GCP2-BLANK-SPIKE



#45 AR-1268-5

R.T.: 7.884 min
Delta R.T.: -0.004 min
Response: 15879797
Conc: 8.99 ng/ml