

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR050923\
 Data File : PR061299.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 May 2023 23:23
 Operator : YP\AJ
 Sample : AR1248ICC1600
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR12485250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 10 06:38:28 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR050923CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed May 10 06:36:04 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.696	2.971	214.8E6	238.5E6	74.696	74.029
2) SA Decachlor...	9.634	8.277	222.1E6	516.6E6	146.496	148.898
Target Compounds						
21) L5 AR-1248-1	4.925	4.102	74798356	97236070	1386.505	1424.728
22) L5 AR-1248-2	5.219	4.352	102.2E6	135.7E6	1344.566	1414.497
23) L5 AR-1248-3	5.432	4.394	122.6E6	139.1E6	1380.258	1401.932
24) L5 AR-1248-4	5.866	4.572	132.3E6	172.0E6	1431.238	1430.255
25) L5 AR-1248-5	5.906	4.987	127.7E6	176.7E6	1426.417	1457.290

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR050923\
Data File : PR061299.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 May 2023 23:23
Operator : YP\AJ
Sample : AR1248ICC1600
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR12485250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 10 06:38:28 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR050923CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed May 10 06:36:04 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

