

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR050923\
 Data File : PR061302.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 May 2023 00:07
 Operator : YP\AJ
 Sample : AR1254ICC400
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR12543203

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 10 06:48:54 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR050923CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed May 10 06:46:32 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.696	2.971	58118151	65716045	20.000	20.000
2) SA Decachlor...	9.636	8.277	64384120	141.8E6	40.000	40.000
Target Compounds						
26) L6 AR-1254-1	5.838	4.945	38564472	74444872	400.000	400.000
27) L6 AR-1254-2	6.076	5.105	58591207	63893230	400.000	400.000
28) L6 AR-1254-3	6.468	5.531	60083812	105.9E6	400.000	400.000
29) L6 AR-1254-4	6.780	5.777	41802805	67207154	400.000	400.000
30) L6 AR-1254-5	7.235	6.226	43322721	98387167	400.000	400.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR050923\
Data File : PR061302.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 May 2023 00:07
Operator : YP\AJ
Sample : AR1254ICC400
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR12543203

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 10 06:48:54 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR050923CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed May 10 06:46:32 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

