

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR051021\
 Data File : PR050234.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 May 2021 17:03
 Operator : DD\AJ
 Sample : AR1254CCC400
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR12543019

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 11 00:36:03 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR050721CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat May 08 06:36:02 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.634	3.816	70209498	104.5E6	20.768	21.455
2) SA Decachlor...	10.504	8.844	226.0E6	276.7E6	41.369	41.828
Target Compounds						
26) L6 AR-1254-1	6.670	5.703	142.2E6	216.7E6	383.464	391.213
27) L6 AR-1254-2	6.888	5.849	225.8E6	186.0E6	380.420	381.217
28) L6 AR-1254-3	7.258	6.252	250.4E6	332.9E6	374.283	380.989
29) L6 AR-1254-4	7.545	6.480	189.6E6	211.9E6	374.048	379.376
30) L6 AR-1254-5	7.964	6.896	204.3E6	301.8E6	373.870	372.248

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR051021\
Data File : PR050234.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 May 2021 17:03
Operator : DD\AJ
Sample : AR1254CCC400
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR12543019

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 11 00:36:03 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR050721CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 08 06:36:02 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

