

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR051021\
 Data File : PR050225.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 May 2021 14:43
 Operator : DD\AJ
 Sample : M2344-10
 Misc :
 ALS Vial : 18 Sample Multiplier: 1



Instrument :
 ECD_R
 ClientSampleId :
 C0B83

Manual Integrations
 APPROVED

Ankita
 5/11/2021 4:14:53 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 11 00:30:36 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR050721CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat May 08 06:36:02 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(16) AR-1242-1 (L4)

R.T.	Response	Conc
5.84	9741904	55.88
5.84	9741904	39.46
5.92	2842162	18.72
6.00	3520785	27.37
6.74	8920967	55.70

(16) AR-1242-1 #2 (L4)

R.T.	Response	Conc
4.92	9897581	53.28
4.92	9897581	36.80
5.10	5151573	35.58
5.18	5675580	40.88
5.70	26465490	133.74

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR051021\
 Data File : PR050225.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 May 2021 14:43
 Operator : DD\AJ
 Sample : M2344-10
 Misc :
 ALS Vial : 18 Sample Multiplier: 1



Instrument :
 ECD_R
 ClientSampleId :
 C0B83

Manual Integrations
 APPROVED

Ankita
 5/11/2021 4:14:53 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 11 00:30:36 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR050721CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat May 08 06:36:02 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(16) AR-1242-1 #2 (L4)

R.T.	Response	Conc
5.82	1279544	7.34
5.84	8138253	32.96
5.92	2842162	18.72
6.00	3520785	27.37
6.74	8920967	55.70

(16) AR-1242-1 #2 (L4)

R.T.	Response	Conc
4.90	1092843	5.88
4.92	8890348	33.05
5.10	5151573	35.58
5.18	5675580	40.88
5.70	26465490	133.74

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR051021\
 Data File : PR050225.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 May 2021 14:43
 Operator : DD\AJ
 Sample : M2344-10
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 C0B83

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 11 00:30:36 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR050721CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat May 08 06:36:02 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Manual Integrations
 APPROVED

Ankita
 5/11/2021 4:14:53 PM

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.633	3.815	75832182	116.4E6	22.431	23.897
2) SA Decachlor...	10.505	8.843	209.3E6	241.2E6	38.312	36.453
Target Compounds						
16) L4 AR-1242-1	5.820	4.905	1279544	1092843	7.340m	5.883m
17) L4 AR-1242-2	5.838	4.922	8138253	8890348	32.962m	33.053m
18) L4 AR-1242-3	5.923	5.098	2842162	5151573	18.719	35.580 #
19) L4 AR-1242-4	6.000	5.180	3520785	5675580	27.374	40.876 #
20) L4 AR-1242-5	6.738	5.703	8920967	26465490	55.699	133.744 #
31) L7 AR-1260-1	7.420	6.382	110.3E6	121.6E6	418.333	365.539
32) L7 AR-1260-2	7.676	6.568	175.8E6	202.1E6	545.221	488.275
33) L7 AR-1260-3	8.035	6.722	88423210	145.7E6	359.162	368.400
34) L7 AR-1260-4	8.269	7.193	124.8E6	111.7E6	426.207	323.345
35) L7 AR-1260-5	8.607	7.432	288.1E6	361.4E6	476.958	416.904

M7
 7/13/21

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.