

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR051319\
 Data File : PR037833.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 May 2019 08:57
 Operator : SM\MA
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 HEXANE

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 14 00:34:30 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR051119.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sun May 12 00:51:44 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds							
7)	L1	AR-1016-5	5.324	0.000	1059149	0	17.077 N.D. #
15)	L3	AR-1232-5	5.324	0.000	1059149	0	30.155 N.D. #
20)	L4	AR-1242-5	5.698	6.634	998247	3031359	16.042 24.505 #
24)	L5	AR-1248-4	5.324	6.634	1059149	3031359	13.143 13.870
25)	L5	AR-1248-5	5.698	6.634	998247	3031359	11.748 14.563
26)	L6	AR-1254-1	5.698	6.634	998247	3031359	6.743 12.586 #
28)	L6	AR-1254-3	6.132	7.223	3123410	3594920	14.149 8.506 #
29)	L6	AR-1254-4	0.000	7.496	0	4742478	N.D. 15.350 #
32)	L7	AR-1260-2	6.528	0.000	990152	0	6.317 N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR051319\
Data File : PR037833.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 May 2019 08:57
Operator : SM\MA
Sample : HEXANE
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampled :
HEXANE

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 14 00:34:30 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR051119.M
Quant Title : GC EXTRACTABLES
QLast Update : Sun May 12 00:51:44 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 0.000 min
Exp R.T. : 3.782 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
HEXANE




#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T. : 4.628 min
Response: 0
Conc: N.D.




#2 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 8.775 min
Response: 0
Conc: N.D.



#2 Decachlorobiphenyl


R.T.: 0.000 min
Exp R.T. : 10.387 min
Response: 0
Conc: N.D.



#7 AR-1016-5


R.T.: 5.324 min
Delta R.T.: 0.021 min
Response: 1059149
Conc: 17.08 ng/ml

Instrument :
ECD_R
ClientSampleId :
HEXANE




#7 AR-1016-5

R.T.: 0.000 min
Exp R.T. : 6.256 min
Response: 0
Conc: N.D.




#15 AR-1232-5

R.T.: 5.324 min
Delta R.T.: 0.021 min
Response: 1059149
Conc: 30.16 ng/ml



#15 AR-1232-5


R.T.: 0.000 min
Exp R.T. : 6.052 min
Response: 0
Conc: N.D.



#20 AR-1242-5


R.T.: 5.698 min
Delta R.T.: 0.047 min
Response: 998247
Conc: 16.04 ng/ml

Instrument :
ECD_R
ClientSampleId :
HEXANE




#20 AR-1242-5

R.T.: 6.634 min
Delta R.T.: -0.059 min
Response: 3031359
Conc: 24.50 ng/ml




#24 AR-1248-4

R.T.: 5.324 min
Delta R.T.: 0.021 min
Response: 1059149
Conc: 13.14 ng/ml



#24 AR-1248-4


R.T.: 6.634 min
Delta R.T.: -0.021 min
Response: 3031359
Conc: 13.87 ng/ml



#25 AR-1248-5


R.T.: 5.698 min
Delta R.T.: 0.006 min
Response: 998247
Conc: 11.75 ng/ml

Instrument :
ECD_R
ClientSampleId :
HEXANE




#25 AR-1248-5

R.T.: 6.634 min
Delta R.T.: -0.059 min
Response: 3031359
Conc: 14.56 ng/ml




#26 AR-1254-1

R.T.: 5.698 min
Delta R.T.: 0.046 min
Response: 998247
Conc: 6.74 ng/ml



#26 AR-1254-1


R.T.: 6.634 min
Delta R.T.: 0.005 min
Response: 3031359
Conc: 12.59 ng/ml



#28 AR-1254-3


R.T.: 6.132 min
Delta R.T.: -0.067 min
Response: 3123410
Conc: 14.15 ng/ml

Instrument :
ECD_R
ClientSampleId :
HEXANE




#28 AR-1254-3

R.T.: 7.223 min
Delta R.T.: 0.015 min
Response: 3594920
Conc: 8.51 ng/ml




#29 AR-1254-4

R.T.: 0.000 min
Exp R.T. : 6.424 min
Response: 0
Conc: N.D.



#29 AR-1254-4


R.T.: 7.496 min
Delta R.T.: 0.006 min
Response: 4742478
Conc: 15.35 ng/ml



#32 AR-1260-2

R.T.: 6.528 min
Delta R.T.: 0.015 min
Response: 990152
Conc: 6.32 ng/ml

Instrument :
ECD_R
ClientSampleId :
HEXANE



#32 AR-1260-2

R.T.: 0.000 min
Exp R.T. : 7.622 min
Response: 0
Conc: N.D.