

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR051421\
 Data File : PR050318.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 May 2021 23:01
 Operator : DD\AJ
 Sample : PB136431BL
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 ABLK431

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 15 01:51:02 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR050721CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat May 08 06:36:02 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.633	3.815	79485622	122.4E6	23.512	25.112
2) SA Decachlor...	10.502	8.840	244.9E6	290.6E6	44.844	43.921
Target Compounds						
9) L2 AR-1221-2	4.944	4.118	1224335	1704624	37.459	40.558
10) L2 AR-1221-3	4.944f	0.000	1224335	0	11.074	N.D. #
11) L3 AR-1232-1	4.944f	0.000	1224335	0	13.130	N.D. #
40) L8 AR-1262-5	9.754f	0.000	1971569	0	6.103	N.D. #
44) L9 AR-1268-4	9.754f	0.000	1971569	0	5.190	N.D. #
45) L9 AR-1268-5	10.153	8.576	1654621	1968786	0.560	0.541



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR051421\
Data File : PR050318.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 May 2021 23:01
Operator : DD\AJ
Sample : PB136431BL
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
ABLK431

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 15 01:51:02 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR050721CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 08 06:36:02 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#9 AR-1221-2


R.T.: 4.944 min
Delta R.T.: 0.010 min
Response: 1224335
Conc: 37.46 ng/ml

Instrument :
ECD_R
ClientSampleId :
ABLK431




#9 AR-1221-2

R.T.: 4.118 min
Delta R.T.: 0.003 min
Response: 1704624
Conc: 40.56 ng/ml




#10 AR-1221-3

R.T.: 4.944 min
Delta R.T.: -0.069 min
Response: 1224335
Conc: 11.07 ng/ml



#10 AR-1221-3


R.T.: 0.000 min
Exp R.T. : 4.191 min
Response: 0
Conc: N.D.



#11 AR-1232-1


R.T.: 4.944 min
Delta R.T.: -0.068 min
Response: 1224335
Conc: 13.13 ng/ml

Instrument :
ECD_R
ClientSampleId :
ABLK431




#11 AR-1232-1

R.T.: 0.000 min
Exp R.T. : 4.190 min
Response: 0
Conc: N.D.




#40 AR-1262-5

R.T.: 9.754 min
Delta R.T.: 0.035 min
Response: 1971569
Conc: 6.10 ng/ml



#40 AR-1262-5


R.T.: 0.000 min
Exp R.T. : 8.277 min
Response: 0
Conc: N.D.



#44 AR-1268-4


R.T.: 9.754 min
Delta R.T.: 0.036 min
Response: 1971569
Conc: 5.19 ng/ml

Instrument :
ECD_R
ClientSampleId :
ABLK431




#44 AR-1268-4

R.T.: 0.000 min
Exp R.T. : 8.277 min
Response: 0
Conc: N.D.



#45 AR-1268-5

R.T.: 10.153 min
Delta R.T.: 0.004 min
Response: 1654621
Conc: 0.56 ng/ml



#45 AR-1268-5

R.T.: 8.576 min
Delta R.T.: -0.001 min
Response: 1968786
Conc: 0.54 ng/ml