

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR051421\
 Data File : PR050338.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 May 2021 04:26
 Operator : DD\AJ
 Sample : AR1254CCC400
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR12543024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 15 06:03:04 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR050721CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat May 08 06:36:02 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.634	3.815	71138162	109.7E6	21.043	22.514
2) SA Decachlor...	10.503	8.839	233.0E6	272.8E6	42.654	41.229
Target Compounds						
26) L6 AR-1254-1	6.670	5.702	145.8E6	219.8E6	393.250	396.796
27) L6 AR-1254-2	6.888	5.848	230.4E6	189.3E6	388.030	388.134
28) L6 AR-1254-3	7.258	6.251	256.0E6	333.7E6	382.689	381.960
29) L6 AR-1254-4	7.545	6.477	191.2E6	212.7E6	377.160	380.793
30) L6 AR-1254-5	7.963	6.894	205.5E6	295.5E6	376.047	364.565

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR051421\
Data File : PR050338.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 May 2021 04:26
Operator : DD\AJ
Sample : AR1254CCC400
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR12543024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 15 06:03:04 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR050721CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 08 06:36:02 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

