

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_R\Data\PR052021\  
 Data File : PR050433.D  
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
 Acq On : 20 May 2021 11:57  
 Operator : DD\AJ  
 Sample : M2408-12DL2 40X  
 Misc :  
 ALS Vial : 11 Sample Multiplier: 1

Instrument :  
 ECD\_R  
 ClientSampleId :  
 C0CLODL2

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: May 21 07:01:06 2021  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_R\Method\PR050721CLP.M  
 Quant Title : GC EXTRACTABLES  
 QLast Update : Sat May 08 06:36:02 2021  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 1 µl  
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2  
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

| Compound | RT#1 | RT#2 | Resp#1 | Resp#2 | ng/ml | ng/ml |
|----------|------|------|--------|--------|-------|-------|
|----------|------|------|--------|--------|-------|-------|

#### System Monitoring Compounds

| Target Compounds |    |           |       |       |         |         |
|------------------|----|-----------|-------|-------|---------|---------|
| 26)              | L6 | AR-1254-1 | 6.669 | 5.701 | 231.9E6 | 505.278 |
| 27)              | L6 | AR-1254-2 | 6.887 | 5.847 | 415.6E6 | 684.327 |
| 28)              | L6 | AR-1254-3 | 7.258 | 6.250 | 515.8E6 | 730.522 |
| 29)              | L6 | AR-1254-4 | 7.544 | 6.476 | 420.5E6 | 808.908 |
| 30)              | L6 | AR-1254-5 | 7.963 | 6.893 | 454.6E6 | 782.374 |

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_R\Data\PR052021\  
Data File : PR050433.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 20 May 2021 11:57  
Operator : DD\AJ  
Sample : M2408-12DL2 40X  
Misc :  
ALS Vial : 11 Sample Multiplier: 1

Instrument :  
ECD\_R  
ClientSampleId :  
C0CL0DL2

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: May 21 07:01:06 2021  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_R\Method\PR050721CLP.M  
Quant Title : GC EXTRACTABLES  
QLast Update : Sat May 08 06:36:02 2021  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 µl  
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2  
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

