

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR052119.M
 Title : GC EXTRACTABLES
 Last Update : Wed May 22 01:39:47 2019
 Response Via : Initial Calibration

Calibration Files

750 =PR038115.D 1000 =PR038114.D 500 =PR038116.D
 250 =PR038117.D 50 =PR038118.D

Compound		750	1000	500	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	1.399	1.320	1.503	1.384	1.500	1.421	E6	5.58
2) SA	Decachlorobiphenyl	2.685	2.574	2.809	2.770	3.213	2.810	E6	8.63
3) L1	AR-1016-1	6.765	6.413	7.560	7.271	8.539	7.309	E4	11.19
4) L1	AR-1016-2	0.953	0.900	1.009	1.001	1.128	0.998	E5	8.50
5) L1	AR-1016-3	5.151	4.850	5.567	5.524	6.034	5.425	E4	8.28
6) L1	AR-1016-4	4.026	3.770	4.406	4.436	4.925	4.312	E4	10.21
7) L1	AR-1016-5	5.608	5.292	6.167	6.138	7.585	6.158	E4	14.27
8) L2	AR-1221-1			1.739			1.739	E4	0.00
9) L2	AR-1221-2			1.276			1.276	E4	0.00
10) L2	AR-1221-3			4.027			4.027	E4	0.00
11) L3	AR-1232-1			3.165			3.165	E4	0.00
12) L3	AR-1232-2			4.107			4.107	E4	0.00
13) L3	AR-1232-3			2.216			2.216	E4	0.00
14) L3	AR-1232-4			1.972			1.972	E4	0.00
15) L3	AR-1232-5			2.319			2.319	E4	0.00
16) L4	AR-1242-1			4.683			4.683	E4	0.00
17) L4	AR-1242-2			6.329			6.329	E4	0.00
18) L4	AR-1242-3			3.561			3.561	E4	0.00
19) L4	AR-1242-4			3.518			3.518	E4	0.00
20) L4	AR-1242-5			5.469			5.469	E4	0.00
21) L5	AR-1248-1			3.766			3.766	E4	0.00
22) L5	AR-1248-2			5.093			5.093	E4	0.00
23) L5	AR-1248-3			5.288			5.288	E4	0.00
24) L5	AR-1248-4			6.868			6.868	E4	0.00
25) L5	AR-1248-5			7.666			7.666	E4	0.00
26) L6	AR-1254-1			1.528			1.528	E5	0.00
27) L6	AR-1254-2			1.347			1.347	E5	0.00
28) L6	AR-1254-3			2.360			2.360	E5	0.00
29) L6	AR-1254-4			1.681			1.681	E5	0.00
30) L6	AR-1254-5			2.267			2.267	E5	0.00
31) L7	AR-1260-1	1.245	1.160	1.328	1.318	1.553	1.321	E5	11.07
32) L7	AR-1260-2	1.673	1.573	1.784	1.751	2.051	1.766	E5	10.11
33) L7	AR-1260-3	1.483	1.398	1.590	1.552	1.809	1.566	E5	9.83
34) L7	AR-1260-4	1.274	1.206	1.361	1.331	1.559	1.346	E5	9.87
35) L7	AR-1260-5	3.712	3.546	3.910	3.739	4.180	3.817	E5	6.29
36) L8	AR-1262-1			8.528			8.528	E4	0.00
37) L8	AR-1262-2			4.333			4.333	E5	0.00
38) L8	AR-1262-3			1.693			1.693	E5	0.00
39) L8	AR-1262-4			3.086			3.086	E5	0.00
40) L8	AR-1262-5			1.423			1.423	E5	0.00
41) L9	AR-1268-1			5.032			5.032	E5	0.00
42) L9	AR-1268-2			4.510			4.510	E5	0.00
43) L9	AR-1268-3			3.756			3.756	E5	0.00
44) L9	AR-1268-4			1.569			1.569	E5	0.00
45) L9	AR-1268-5			1.152			1.152	E6	0.00

Signal #2 Calibration Files

750 =PR038115.D 1000 =PR038114.D 500 =PR038116.D
 250 =PR038117.D 50 =PR038118.D

Compound		750	1000	500	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	4.358	4.080	4.709	4.398	4.616	4.432	E6	5.54
2) SA	Decachlorobiphenyl	6.238	5.973	6.495	6.444	7.300	6.490	E6	7.66
3) L1	AR-1016-1	1.818	1.745	2.004	1.944	2.207	1.944	E5	9.23
4) L1	AR-1016-2	2.697	2.522	2.903	2.785	3.098	2.801	E5	7.72
5) L1	AR-1016-3	1.626	1.532	1.772	1.729	1.974	1.727	E5	9.68

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR052119.M
 Title : GC EXTRACTABLES
 Last Update : Wed May 22 01:39:47 2019
 Response Via : Initial Calibration

Calibration Files

750 =PR038115.D 1000 =PR038114.D 500 =PR038116.D
 250 =PR038117.D 50 =PR038118.D

Compound			750	1000	500	250	50	Avg		%RSD
6)	L1	AR-1016-4	1.378	1.300	1.491	1.444	1.578	1.438	E5	7.39
7)	L1	AR-1016-5	1.396	1.312	1.525	1.480	1.738	1.490	E5	10.79
8)	L2	AR-1221-1			4.679			4.679	E4	0.00
9)	L2	AR-1221-2			3.553			3.553	E4	0.00
10)	L2	AR-1221-3			1.154			1.154	E5	0.00
11)	L3	AR-1232-1			9.047			9.047	E4	0.00
12)	L3	AR-1232-2			5.148			5.148	E4	0.00
13)	L3	AR-1232-3			1.124			1.124	E5	0.00
14)	L3	AR-1232-4			5.780			5.780	E4	0.00
15)	L3	AR-1232-5			4.204			4.204	E4	0.00
16)	L4	AR-1242-1			1.215			1.215	E5	0.00
17)	L4	AR-1242-2			1.719			1.719	E5	0.00
18)	L4	AR-1242-3			1.090			1.090	E5	0.00
19)	L4	AR-1242-4			9.279			9.279	E4	0.00
20)	L4	AR-1242-5			1.155			1.155	E5	0.00
21)	L5	AR-1248-1			9.621			9.621	E4	0.00
22)	L5	AR-1248-2			1.375			1.375	E5	0.00
23)	L5	AR-1248-3			1.658			1.658	E5	0.00
24)	L5	AR-1248-4			2.070			2.070	E5	0.00
25)	L5	AR-1248-5			1.978			1.978	E5	0.00
26)	L6	AR-1254-1			2.698			2.698	E5	0.00
27)	L6	AR-1254-2			4.319			4.319	E5	0.00
28)	L6	AR-1254-3			5.000			5.000	E5	0.00
29)	L6	AR-1254-4			3.769			3.769	E5	0.00
30)	L6	AR-1254-5			4.181			4.181	E5	0.00
31)	L7	AR-1260-1	2.816	2.660	3.030	2.976	3.455	2.987	E5	10.00
32)	L7	AR-1260-2	3.480	3.290	3.722	3.641	4.088	3.644	E5	8.18
33)	L7	AR-1260-3	2.673	2.534	2.859	2.809	3.203	2.816	E5	8.91
34)	L7	AR-1260-4	3.232	3.067	3.430	3.319	3.748	3.359	E5	7.58
35)	L7	AR-1260-5	6.853	6.553	7.204	6.867	7.474	6.990	E5	5.08
36)	L8	AR-1262-1			4.124			4.124	E5	0.00
37)	L8	AR-1262-2			8.061			8.061	E5	0.00
38)	L8	AR-1262-3			5.297			5.297	E5	0.00
39)	L8	AR-1262-4			3.870			3.870	E5	0.00
40)	L8	AR-1262-5			2.861			2.861	E5	0.00
41)	L9	AR-1268-1			8.888			8.888	E5	0.00
42)	L9	AR-1268-2			8.177			8.177	E5	0.00
43)	L9	AR-1268-3			6.737			6.737	E5	0.00
44)	L9	AR-1268-4			3.116			3.116	E5	0.00
45)	L9	AR-1268-5			2.365			2.365	E6	0.00

(#) = Out of Range