

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR052221.M
 Title : GC EXTRACTABLES
 Last Update : Sat May 22 08:50:48 2021
 Response Via : Initial Calibration

Calibration Files

750 =PR050473.D 1000 =PR050472.D 500 =PR050474.D
 250 =PR050475.D 50 =PR050476.D

Compound		750	1000	500	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	4.212	4.197	4.264	4.140	4.137	4.190	E6	1.27
2) SA	Decachlorobiphenyl	6.438	6.488	6.723	6.783	7.371	6.761	E6	5.50
3) L1	AR-1016-1	2.208	2.203	2.300	2.338	2.512	2.312	E5	5.45
4) L1	AR-1016-2	3.163	3.197	3.342	3.298	3.476	3.295	E5	3.78
5) L1	AR-1016-3	1.869	1.863	1.979	2.016	2.174	1.980	E5	6.45
6) L1	AR-1016-4	1.613	1.615	1.687	1.708	1.751	1.675	E5	3.60
7) L1	AR-1016-5	1.628	1.621	1.735	1.763	1.900	1.729	E5	6.61
8) L2	AR-1221-1			5.176			5.176	E4	0.00
9) L2	AR-1221-2			4.028			4.028	E4	0.00
10) L2	AR-1221-3			1.244			1.244	E5	0.00
11) L3	AR-1232-1			9.646			9.646	E4	0.00
12) L3	AR-1232-2			6.111			6.111	E4	0.00
13) L3	AR-1232-3			1.323			1.323	E5	0.00
14) L3	AR-1232-4			6.705			6.705	E4	0.00
15) L3	AR-1232-5			5.280			5.280	E4	0.00
16) L4	AR-1242-1			1.703			1.703	E5	0.00
17) L4	AR-1242-2			2.454			2.454	E5	0.00
18) L4	AR-1242-3			1.469			1.469	E5	0.00
19) L4	AR-1242-4			1.249			1.249	E5	0.00
20) L4	AR-1242-5			1.458			1.458	E5	0.00
21) L5	AR-1248-1			1.305			1.305	E5	0.00
22) L5	AR-1248-2			1.887			1.887	E5	0.00
23) L5	AR-1248-3			2.173			2.173	E5	0.00
24) L5	AR-1248-4			2.632			2.632	E5	0.00
25) L5	AR-1248-5			2.508			2.508	E5	0.00
26) L6	AR-1254-1			2.619			2.619	E5	0.00
27) L6	AR-1254-2			4.162			4.162	E5	0.00
28) L6	AR-1254-3			4.556			4.556	E5	0.00
29) L6	AR-1254-4			3.493			3.493	E5	0.00
30) L6	AR-1254-5			3.886			3.886	E5	0.00
31) L7	AR-1260-1	3.178	3.223	3.370	3.369	3.733	3.375	E5	6.47
32) L7	AR-1260-2	3.932	4.000	4.176	4.196	4.540	4.169	E5	5.67
33) L7	AR-1260-3	2.901	2.934	3.061	3.111	3.356	3.073	E5	5.88
34) L7	AR-1260-4	3.478	3.535	3.650	3.669	3.879	3.642	E5	4.24
35) L7	AR-1260-5	7.365	7.517	7.644	7.454	7.511	7.498	E5	1.36
36) L8	AR-1262-1			4.384			4.384	E5	0.00
37) L8	AR-1262-2			8.409			8.409	E5	0.00
38) L8	AR-1262-3			5.429			5.429	E5	0.00
39) L8	AR-1262-4			3.982			3.982	E5	0.00
40) L8	AR-1262-5			2.924			2.924	E5	0.00
41) L9	AR-1268-1			9.555			9.555	E5	0.00
42) L9	AR-1268-2			8.586			8.586	E5	0.00
43) L9	AR-1268-3			7.235			7.235	E5	0.00
44) L9	AR-1268-4			3.180			3.180	E5	0.00
45) L9	AR-1268-5			2.544			2.544	E6	0.00

Signal #2 Calibration Files

750 =PR050473.D 1000 =PR050472.D 500 =PR050474.D
 250 =PR050475.D 50 =PR050476.D

Compound		750	1000	500	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	6.038	6.100	6.135	5.791	5.463	5.905	E6	4.77
2) SA	Decachlorobiphenyl	7.295	7.386	7.583	7.504	7.891	7.532	E6	3.04
3) L1	AR-1016-1	2.252	2.271	2.334	2.308	2.431	2.319	E5	3.03
4) L1	AR-1016-2	3.285	3.304	3.387	3.342	3.321	3.328	E5	1.18
5) L1	AR-1016-3	1.708	1.713	1.770	1.776	1.821	1.757	E5	2.69

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR052221.M
 Title : GC EXTRACTABLES
 Last Update : Sat May 22 08:50:48 2021
 Response Via : Initial Calibration

Calibration Files

750 =PR050473.D 1000 =PR050472.D 500 =PR050474.D
 250 =PR050475.D 50 =PR050476.D

	Compound	750	1000	500	250	50	Avg		%RSD
6)	L1 AR-1016-4	1.284	1.280	1.353	1.407	1.533	1.372	E5	7.60
7)	L1 AR-1016-5	1.729	1.747	1.844	1.865	2.029	1.843	E5	6.50
8)	L2 AR-1221-1			6.374			6.374	E4	0.00
9)	L2 AR-1221-2			4.913			4.913	E4	0.00
10)	L2 AR-1221-3			1.554			1.554	E5	0.00
11)	L3 AR-1232-1			1.184			1.184	E5	0.00
12)	L3 AR-1232-2			1.333			1.333	E5	0.00
13)	L3 AR-1232-3			7.013			7.013	E4	0.00
14)	L3 AR-1232-4			6.155			6.155	E4	0.00
15)	L3 AR-1232-5			6.990			6.990	E4	0.00
16)	L4 AR-1242-1			1.712			1.712	E5	0.00
17)	L4 AR-1242-2			2.472			2.472	E5	0.00
18)	L4 AR-1242-3			1.310			1.310	E5	0.00
19)	L4 AR-1242-4			1.263			1.263	E5	0.00
20)	L4 AR-1242-5			1.811			1.811	E5	0.00
21)	L5 AR-1248-1			1.308			1.308	E5	0.00
22)	L5 AR-1248-2			1.759			1.759	E5	0.00
23)	L5 AR-1248-3			1.856			1.856	E5	0.00
24)	L5 AR-1248-4			2.324			2.324	E5	0.00
25)	L5 AR-1248-5			2.495			2.495	E5	0.00
26)	L6 AR-1254-1			3.757			3.757	E5	0.00
27)	L6 AR-1254-2			3.359			3.359	E5	0.00
28)	L6 AR-1254-3			5.856			5.856	E5	0.00
29)	L6 AR-1254-4			3.747			3.747	E5	0.00
30)	L6 AR-1254-5			5.534			5.534	E5	0.00
31)	L7 AR-1260-1	3.654	3.686	3.838	3.857	4.148	3.837	E5	5.11
32)	L7 AR-1260-2	4.572	4.606	4.819	4.811	5.133	4.788	E5	4.67
33)	L7 AR-1260-3	4.389	4.444	4.570	4.577	4.709	4.538	E5	2.77
34)	L7 AR-1260-4	3.704	3.758	3.875	3.837	4.009	3.837	E5	3.05
35)	L7 AR-1260-5	9.599	9.710	9.969	9.763	9.641	9.736	E5	1.48
36)	L8 AR-1262-1			2.779			2.779	E5	0.00
37)	L8 AR-1262-2			1.051			1.051	E6	0.00
38)	L8 AR-1262-3			3.793			3.793	E5	0.00
39)	L8 AR-1262-4			7.596			7.596	E5	0.00
40)	L8 AR-1262-5			3.430			3.430	E5	0.00
41)	L9 AR-1268-1			1.141			1.141	E6	0.00
42)	L9 AR-1268-2			1.059			1.059	E6	0.00
43)	L9 AR-1268-3			8.760			8.760	E5	0.00
44)	L9 AR-1268-4			3.743			3.743	E5	0.00
45)	L9 AR-1268-5			2.892			2.892	E6	0.00

(#) = Out of Range