

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR052221\
 Data File : PR050472.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 May 2021 05:33
 Operator : DD\AJ
 Sample : AR1660ICC1000
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR1660ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 22 07:06:56 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR052221.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat May 22 07:06:36 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.632	3.814	419.7E6	610.0E6	98.432	99.429
2)	SA Decachlor...	10.496	8.835	648.8E6	738.6E6	96.502	97.398
Target Compounds							
3)	L1 AR-1016-1	5.813	4.899	220.3E6	227.1E6	957.618	972.898
4)	L1 AR-1016-2	5.836	4.919	319.7E6	330.4E6	956.463	975.578
5)	L1 AR-1016-3	5.898	5.095	186.3E6	171.3E6	941.445	967.615
6)	L1 AR-1016-4	5.999	5.135	161.5E6	128.0E6	957.515	946.042
7)	L1 AR-1016-5	6.292	5.349	162.1E6	174.7E6	933.939	947.248
31)	L7 AR-1260-1	7.416	6.378	322.3E6	368.6E6	956.334	960.466
32)	L7 AR-1260-2	7.673	6.564	400.0E6	460.6E6	957.780	955.880
33)	L7 AR-1260-3	8.031	6.717	293.4E6	444.4E6	958.624	972.319
34)	L7 AR-1260-4	8.266	7.188	353.5E6	375.8E6	968.587	969.635
35)	L7 AR-1260-5	8.601	7.427	751.7E6	971.0E6	983.367	973.986



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR052221\
Data File : PR050472.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 May 2021 05:33
Operator : DD\AJ
Sample : AR1660ICC1000
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR1660ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 22 07:06:56 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR052221.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 22 07:06:36 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.632 min
Delta R.T.: 0.000 min
Response: 419697981
Conc: 98.43 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1660ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.814 min
Delta R.T.: 0.000 min
Response: 609964489
Conc: 99.43 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.496 min
Delta R.T.: 0.000 min
Response: 648802333
Conc: 96.50 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.835 min
Delta R.T.: 0.000 min
Response: 738614626
Conc: 97.40 ng/ml



#3 AR-1016-1


R.T.: 5.813 min
Delta R.T.: 0.000 min
Response: 220256486
Conc: 957.62 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1660ICC1000




#3 AR-1016-1

R.T.: 4.899 min
Delta R.T.: 0.000 min
Response: 227103725
Conc: 972.90 ng/ml




#4 AR-1016-2

R.T.: 5.836 min
Delta R.T.: 0.000 min
Response: 319684419
Conc: 956.46 ng/ml



#4 AR-1016-2


R.T.: 4.919 min
Delta R.T.: 0.000 min
Response: 330418776
Conc: 975.58 ng/ml



#5 AR-1016-3


R.T.: 5.898 min
Delta R.T.: 0.000 min
Response: 186274026
Conc: 941.45 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1660ICC1000




#5 AR-1016-3

R.T.: 5.095 min
Delta R.T.: 0.000 min
Response: 171251674
Conc: 967.61 ng/ml




#6 AR-1016-4

R.T.: 5.999 min
Delta R.T.: 0.000 min
Response: 161486953
Conc: 957.52 ng/ml



#6 AR-1016-4


R.T.: 5.135 min
Delta R.T.: 0.000 min
Response: 128043238
Conc: 946.04 ng/ml



#7 AR-1016-5


R.T.: 6.292 min
Delta R.T.: 0.000 min
Response: 162055904
Conc: 933.94 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1660ICC1000




#7 AR-1016-5

R.T.: 5.349 min
Delta R.T.: 0.000 min
Response: 174689679
Conc: 947.25 ng/ml




#31 AR-1260-1

R.T.: 7.416 min
Delta R.T.: 0.000 min
Response: 322297952
Conc: 956.33 ng/ml



#31 AR-1260-1


R.T.: 6.378 min
Delta R.T.: 0.000 min
Response: 368610121
Conc: 960.47 ng/ml



#32 AR-1260-2


R.T.: 7.673 min
Delta R.T.: 0.000 min
Response: 399963740
Conc: 957.78 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1660ICC1000




#32 AR-1260-2

R.T.: 6.564 min
Delta R.T.: 0.000 min
Response: 460619753
Conc: 955.88 ng/ml




#33 AR-1260-3

R.T.: 8.031 min
Delta R.T.: 0.000 min
Response: 293411444
Conc: 958.62 ng/ml



#33 AR-1260-3


R.T.: 6.717 min
Delta R.T.: 0.000 min
Response: 444358832
Conc: 972.32 ng/ml



#34 AR-1260-4


R.T.: 8.266 min
Delta R.T.: 0.000 min
Response: 353527066
Conc: 968.59 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1660ICC1000




#34 AR-1260-4

R.T.: 7.188 min
Delta R.T.: 0.000 min
Response: 375770751
Conc: 969.63 ng/ml



#35 AR-1260-5

R.T.: 8.601 min
Delta R.T.: 0.000 min
Response: 751699196
Conc: 983.37 ng/ml



#35 AR-1260-5

R.T.: 7.427 min
Delta R.T.: 0.000 min
Response: 970952719
Conc: 973.99 ng/ml