

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR052522\
 Data File : PR054364.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 May 2022 03:51
 Operator : AJ\MA
 Sample : N2897-06
 Misc :
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 P042-SS001-0006-01

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/27/2022
 Supervised By :Ankita Jodhani 05/27/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 26 04:33:58 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR052522.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed May 25 15:27:54 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.649	2.922	84654175	43422156	22.951	22.988
2) SA Decachlor...	9.527	8.158	24614066	32664248	16.111m	22.682 #

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR052522\
Data File : PR054364.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 May 2022 03:51
Operator : AJ\MA
Sample : N2897-06
Misc :
ALS Vial : 41 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
P042-SS001-0006-01


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/27/2022
Supervised By :Ankita Jodhani 05/27/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 26 04:33:58 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR052522.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed May 25 15:27:54 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.649 min
Delta R.T.: 0.000 min
Response: 84654175
Conc: 22.95 ng/ml

Instrument :
ECD_R
ClientSampleId :
P042-SS001-0006-01


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/27/2022
Supervised By :Ankita Jodhani 05/27/2022




#1 Tetrachloro-m-xylene

R.T.: 2.922 min
Delta R.T.: 0.000 min
Response: 43422156
Conc: 22.99 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.527 min
Delta R.T.: -0.012 min
Response: 24614066
Conc: 16.11 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.158 min
Delta R.T.: -0.004 min
Response: 32664248
Conc: 22.68 ng/ml