

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR052622\
 Data File : PR054432.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 May 2022 22:38
 Operator : AJ\MA
 Sample : N2923-14
 Misc :
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 P024-SS007-0612-01

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 27 01:59:00 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR052522.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed May 25 15:27:54 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.649	2.921	85605415	44187376	23.209	23.393
2) SA Decachlor...	9.531	8.158	31300798	30840881	20.488	21.416
Target Compounds						
31) L7 AR-1260-1	6.583	5.583	60719075	51836267	528.425	633.671
32) L7 AR-1260-2	6.865	5.787	170.8E6	171.0E6	1313.094	1732.916 #
33) L7 AR-1260-3	7.254	5.946	146.1E6	84773509	1567.328	928.452 #
34) L7 AR-1260-4	7.495	6.449	141.6E6	126.1E6	1345.894	1643.374
35) L7 AR-1260-5	7.834	6.713	317.4E6	439.0E6	1686.649	2462.351 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR052622\
Data File : PR054432.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 May 2022 22:38
Operator : AJ\MA
Sample : N2923-14
Misc :
ALS Vial : 46 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
P024-SS007-0612-01

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 27 01:59:00 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR052522.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed May 25 15:27:54 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm

