

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR052622\
 Data File : PR054403.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 May 2022 14:18
 Operator : AJ\MA
 Sample : N2921-02DL 2X
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 P013-SS002-0612-01DL

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/27/2022
 Supervised By :Ankita Jodhani 05/27/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 27 01:33:15 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR052522.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed May 25 15:27:54 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.649	2.921	40508263	20953892	10.982	11.093
2) SA Decachlor...	9.530	8.159	10842108	15127899	7.097m	10.505 #
Target Compounds						
31) L7 AR-1260-1	6.582	5.583	36812791	25819730	320.374m	315.633
32) L7 AR-1260-2	6.864	5.786	50100876	46353531	385.218	469.832
33) L7 AR-1260-3	7.254	5.946	69403480	35989488	744.419	394.162m#
34) L7 AR-1260-4	7.495	6.450	61352891	54524109	583.045	710.367
35) L7 AR-1260-5	7.832	6.714	127.9E6	169.0E6	679.442	947.573 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR052622\
Data File : PR054403.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 May 2022 14:18
Operator : AJ\MA
Sample : N2921-02DL 2X
Misc :
ALS Vial : 21 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
P013-SS002-0612-01DL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/27/2022
Supervised By :Ankita Jodhani 05/27/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 27 01:33:15 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR052522.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed May 25 15:27:54 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.649 min
Delta R.T.: 0.000 min
Response: 40508263
Conc: 10.98 ng/ml

Instrument :
ECD_R
ClientSampleId :
P013-SS002-0612-01DL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/27/2022
Supervised By :Ankita Jodhani 05/27/2022




#1 Tetrachloro-m-xylene

R.T.: 2.921 min
Delta R.T.: 0.000 min
Response: 20953892
Conc: 11.09 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.530 min
Delta R.T.: -0.009 min
Response: 10842108
Conc: 7.10 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.159 min
Delta R.T.: -0.003 min
Response: 15127899
Conc: 10.50 ng/ml




#31 AR-1260-1

R.T.: 6.582 min
Delta R.T.: -0.003 min
Response: 36812791
Conc: 320.37 ng/ml

Instrument :
ECD_R
ClientSampleId :
P013-SS002-0612-01DL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/27/2022
Supervised By :Ankita Jodhani 05/27/2022




#31 AR-1260-1

R.T.: 5.583 min
Delta R.T.: -0.003 min
Response: 25819730
Conc: 315.63 ng/ml




#32 AR-1260-2

R.T.: 6.864 min
Delta R.T.: -0.003 min
Response: 50100876
Conc: 385.22 ng/ml



#32 AR-1260-2

R.T.: 5.786 min
Delta R.T.: -0.003 min
Response: 46353531
Conc: 469.83 ng/ml




#33 AR-1260-3

R.T.: 7.254 min
Delta R.T.: -0.004 min
Response: 69403480
Conc: 744.42 ng/ml

Instrument :
ECD_R
ClientSampleId :
P013-SS002-0612-01DL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/27/2022
Supervised By :Ankita Jodhani 05/27/2022




#33 AR-1260-3

R.T.: 5.946 min
Delta R.T.: -0.001 min
Response: 35989488
Conc: 394.16 ng/ml m




#34 AR-1260-4

R.T.: 7.495 min
Delta R.T.: -0.004 min
Response: 61352891
Conc: 583.05 ng/ml



#34 AR-1260-4

R.T.: 6.450 min
Delta R.T.: -0.002 min
Response: 54524109
Conc: 710.37 ng/ml




#35 AR-1260-5

R.T.: 7.832 min
Delta R.T.: -0.003 min
Response: 127868805
Conc: 679.44 ng/ml

Instrument :
ECD_R
ClientSampleId :
P013-SS002-0612-01DL

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/27/2022
Supervised By :Ankita Jodhani 05/27/2022



#35 AR-1260-5

R.T.: 6.714 min
Delta R.T.: -0.002 min
Response: 168955905
Conc: 947.57 ng/ml