

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR052624\
Data File : PR067023.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 May 2024 15:59
Operator : AJ\MA
Sample : P2590-15MS
Misc :
ALS Vial : 24 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
BHBM0MS

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 05/28/2024
Supervised By :Ankita Jodhani 05/28/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 27 02:26:36 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR052124CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed May 22 04:28:16 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(1) Tetrachloro-m-xylene (SA)

3.650min 15.408 ng/ml

response 94519268

(1) Tetrachloro-m-xylene #2 (SA)

2.892min 18.838 ng/ml

response 87709831

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR052624\
Data File : PR067023.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 May 2024 15:59
Operator : AJ\MA
Sample : P2590-15MS
Misc :
ALS Vial : 24 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
BHBM0MS

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 05/28/2024
Supervised By :Ankita Jodhani 05/28/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 27 02:26:36 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR052124CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed May 22 04:28:16 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(1) Tetrachloro-m-xylene (SA)

3.649min 13.219 ng/ml m

response 81090375

(1) Tetrachloro-m-xylene #2 (SA)

2.891min 15.767 ng/ml m

response 73410350

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR052624\
 Data File : PR067023.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 May 2024 15:59
 Operator : AJ\MA
 Sample : P2590-15MS
 Misc :
 ALS Vial : 24 Sample Multiplier: 1


Instrument :
 ECD_R
 ClientSampleId :
 BHBM0MS

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 05/28/2024
 Supervised By :Ankita Jodhani 05/28/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 27 02:26:36 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR052124CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed May 22 04:28:16 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(2) Decachlorobiphenyl (SA)

9.526min 191.077 ng/ml

response 248107449

(2) Decachlorobiphenyl #2 (SA)

8.115min 145.943 ng/ml

response 324935239

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR052624\
Data File : PR067023.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 May 2024 15:59
Operator : AJ\MA
Sample : P2590-15MS
Misc :
ALS Vial : 24 Sample Multiplier: 1


Instrument :
ECD_R
ClientSampleId :
BHBM0MS

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 05/28/2024
Supervised By :Ankita Jodhani 05/28/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 27 02:26:36 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR052124CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed May 22 04:28:16 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(2) Decachlorobiphenyl (SA)

9.525min 194.562 ng/ml m

response 252631992

(2) Decachlorobiphenyl #2 (SA)

8.115min 148.307 ng/ml m

response 330198958

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR052624\
 Data File : PR067023.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 May 2024 15:59
 Operator : AJ\MA
 Sample : P2590-15MS
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :

ECD_R

ClientSampleId :

BHBM0MS



Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 05/28/2024

Supervised By :Ankita Jodhani 05/28/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 27 02:26:36 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR052124CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed May 22 04:28:16 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(3) AR-1016-1 (L1)
 R.T. Response Conc
 4.87 80002319 498.78
 4.89 114763673 446.99
 4.95 92947636 539.97
 5.06 61584579 464.31
 5.37 45748717 338.97
 (3) AR-1016-1 #2 (L1)
 R.T. Response Conc
 4.00 58992177 392.97
 4.02 72172447 325.71
 4.20 86890331 705.26
 4.24 31741641 303.92
 4.46 62146178 469.22

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR052624\
Data File : PR067023.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 May 2024 15:59
Operator : AJ\MA
Sample : P2590-15MS
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :

ECD_R

ClientSampleId :


BHBM0MS

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 05/28/2024
Supervised By :Ankita Jodhani 05/28/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 27 02:26:36 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR052124CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed May 22 04:28:16 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(41) AR-1268-1 (L9)
R.T. Response Conc
8.14 94417821 227.59
8.22 31016346 83.71
8.44 109357898 336.14
8.84 29375472 258.66
9.22 186140759 220.99
(41) AR-1268-1 #2 (L9)
R.T. Response Conc
6.98 62942897 94.59
7.04 140562462 230.63
7.27 117854840 220.35
7.58 41154542 186.72
7.88 256862073 172.20

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR052624\
 Data File : PR067023.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 May 2024 15:59
 Operator : AJ\MA
 Sample : P2590-15MS
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :

ECD_R

ClientSampleId :


BHBM0MS

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 05/28/2024
 Supervised By :Ankita Jodhani 05/28/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 27 02:26:36 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR052124CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed May 22 04:28:16 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(41) AR-1268-1 (L9)
 R.T. Response Conc
 8.14 94417821 227.59
 8.22 56225227 151.75
 8.44 109357898 336.14
 8.84 29375472 258.66
 9.22 186140759 220.99
 (41) AR-1268-1 #2 (L9)
 R.T. Response Conc
 6.98 62942897 94.59
 7.04 140562462 230.63
 7.27 117854840 220.35
 7.58 41154542 186.72
 7.88 256862073 172.20