

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR052722\
 Data File : PR054474.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 May 2022 11:08
 Operator : AJ\MA
 Sample : N2931-03
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 P009-SS003-1218-01

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 27 23:30:43 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR052522.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed May 25 15:27:54 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.647	2.920	87152993	43904493	23.629	23.243
2) SA Decachlor...	9.538	8.158	25593330	24333738	16.752	16.897
Target Compounds						
31) L7 AR-1260-1	6.584	5.581	57886978	54271209	503.778	663.437 #
32) L7 AR-1260-2	6.867	5.785	110.0E6	110.4E6	845.656	1119.440 #
33) L7 AR-1260-3	7.258	5.943	112.1E6	79525466	1202.730	870.975 #
34) L7 AR-1260-4	7.499	6.448	118.6E6	108.1E6	1126.665	1408.035
35) L7 AR-1260-5	7.837	6.712	232.7E6	290.0E6	1236.599	1626.423 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR052722\
Data File : PR054474.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 May 2022 11:08
Operator : AJ\MA
Sample : N2931-03
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
P009-SS003-1218-01

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 27 23:30:43 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR052522.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed May 25 15:27:54 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm

