

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR053023\
 Data File : PR061614.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 May 2023 16:47
 Operator : YP\AJ
 Sample : AR1660CCC400
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR16603343

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/31/2023
 Supervised By :Ankita Jodhani 05/31/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 30 21:18:55 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR050923CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed May 10 07:27:40 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.694	2.965	47992257	56162002	17.137	18.795
2)	SA Decachlor...	9.631	8.258	55576817	142.2E6	40.218	44.263
Target Compounds							
3)	L1 AR-1016-1	4.923	4.092	29938808	39113229	366.798	375.580m
4)	L1 AR-1016-2	4.946	4.110	43319781	56057638	358.407	378.685
5)	L1 AR-1016-3	5.010	4.291	28011972	31122585	369.958	393.634
6)	L1 AR-1016-4	5.113	4.342	22246449	24753836	364.521	407.538
7)	L1 AR-1016-5	5.430	4.561	22652096	32561209	392.854	393.134
31)	L7 AR-1260-1	6.646	5.659	40975913	69134973	396.793	405.403
32)	L7 AR-1260-2	6.929	5.864	43720137	84254096	371.451	402.706
33)	L7 AR-1260-3	7.319	6.025	31342271	80827783	377.767	389.228
34)	L7 AR-1260-4	7.562	6.534	34876301	70528058	375.390	417.233
35)	L7 AR-1260-5	7.900	6.798	62491818	165.8E6	353.302	409.033

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR053023\
Data File : PR061614.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 May 2023 16:47
Operator : YP\AJ
Sample : AR1660CCC400
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR16603343

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/31/2023
Supervised By :Ankita Jodhani 05/31/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 30 21:18:55 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR050923CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed May 10 07:27:40 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

