

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR053122\
 Data File : PR054567.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 May 2022 18:14
 Operator : AJ\MA
 Sample : AR1248ICC500
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR1248ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 01 05:00:55 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR053122.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 01 05:00:48 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.650	2.924	200.5E6	92783550	50.000	50.000
2) SA Decachlor...	9.534	8.156	80014984	81576787	50.000	50.000
Target Compounds						
21) L5 AR-1248-1	4.868	4.033	35671463	18463589	500.000	500.000
22) L5 AR-1248-2	5.160	4.280	44777733	24411935	500.000	500.000
23) L5 AR-1248-3	5.372	4.321	50001332	25189059	500.000	500.000
24) L5 AR-1248-4	5.804	4.495	49397856	30600201	500.000	500.000
25) L5 AR-1248-5	5.844	4.905	47853116	31085527	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR053122\
Data File : PR054567.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 May 2022 18:14
Operator : AJ\MA
Sample : AR1248ICC500
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR1248ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 01 05:00:55 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR053122.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 01 05:00:48 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

