

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR060121\
 Data File : PR050648.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Jun 2021 17:58
 Operator : DD\AJ
 Sample : AR1248ICC1600
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR12485001

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 02 02:12:10 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR060121CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 02 02:10:19 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.631	3.813	476.4E6	495.1E6	80.100	80.566
2) SA Decachlor...	10.503	8.840	1139.3E6	1228.2E6	147.528	147.546
Target Compounds						
21) L5 AR-1248-1	5.813	4.900	232.2E6	217.3E6	1517.795	1518.352
22) L5 AR-1248-2	6.085	5.135	326.1E6	299.7E6	1450.680	1469.124
23) L5 AR-1248-3	6.292	5.178	382.5E6	318.6E6	1488.751	1467.846
24) L5 AR-1248-4	6.698	5.349	469.9E6	400.3E6	1505.884	1490.965
25) L5 AR-1248-5	6.737	5.740	454.1E6	435.3E6	1520.573	1512.772

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR060121\
Data File : PR050648.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Jun 2021 17:58
Operator : DD\AJ
Sample : AR1248ICC1600
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampled :
AR12485001

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 02 02:12:10 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR060121CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 02 02:10:19 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

