

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR060623CLP.M
 Title : GC EXTRACTABLES
 Last Update : Wed Jun 07 03:46:24 2023
 Response Via : Initial Calibration

Calibration Files

400 =PR061666.D 1600 =PR061668.D 800 =PR061667.D
 200 =PR061665.D 100 =PR061664.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	2.385	2.220	2.339	2.502	2.531	2.395	E6	5.28
2) SA	Decachlorobiphenyl	1.394	1.175	1.304	1.466	1.538	1.375	E6	10.29
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1	4.588	3.960	4.361	5.000	5.102	4.602	E4	10.18
22) L5	AR-1248-2	6.706	5.510	6.186	7.236	7.856	6.699	E4	13.56
23) L5	AR-1248-3	7.660	6.457	7.128	8.172	8.521	7.587	E4	10.84
24) L5	AR-1248-4	7.704	6.615	7.337	8.231	8.805	7.738	E4	10.82
25) L5	AR-1248-5	7.471	6.404	7.085	7.954	8.295	7.442	E4	9.96
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PR061666.D 1600 =PR061668.D 800 =PR061667.D
 200 =PR061665.D 100 =PR061664.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	3.019	2.761	2.957	3.161	3.346	3.049	E6	7.20
2) SA	Decachlorobiphenyl	3.661	3.293	3.561	3.863	4.082	3.692	E6	8.11
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR060623CLP.M
 Title : GC EXTRACTABLES
 Last Update : Wed Jun 07 03:46:24 2023
 Response Via : Initial Calibration

Calibration Files

400 =PR061666.D 1600 =PR061668.D 800 =PR061667.D
 200 =PR061665.D 100 =PR061664.D

Compound			400	1600	800	200	100	Avg	%RSD

6) L1	AR-1016-4							0.000	-1.00
7) L1	AR-1016-5							0.000	-1.00
8) L2	AR-1221-1							0.000	-1.00
9) L2	AR-1221-2							0.000	-1.00
10) L2	AR-1221-3							0.000	-1.00
11) L3	AR-1232-1							0.000	-1.00
12) L3	AR-1232-2							0.000	-1.00
13) L3	AR-1232-3							0.000	-1.00
14) L3	AR-1232-4							0.000	-1.00
15) L3	AR-1232-5							0.000	-1.00
16) L4	AR-1242-1							0.000	-1.00
17) L4	AR-1242-2							0.000	-1.00
18) L4	AR-1242-3							0.000	-1.00
19) L4	AR-1242-4							0.000	-1.00
20) L4	AR-1242-5							0.000	-1.00
21) L5	AR-1248-1		6.464	5.434	6.104	7.000	7.668	6.534 E4	13.04
22) L5	AR-1248-2		0.917	0.755	0.852	1.002	1.099	0.925 E5	14.36
23) L5	AR-1248-3		0.946	0.783	0.883	1.018	1.115	0.949 E5	13.36
24) L5	AR-1248-4		1.134	0.960	1.073	1.219	1.320	1.141 E5	12.04
25) L5	AR-1248-5		1.130	0.976	1.084	1.223	1.291	1.141 E5	10.72
26) L6	AR-1254-1							0.000	-1.00
27) L6	AR-1254-2							0.000	-1.00
28) L6	AR-1254-3							0.000	-1.00
29) L6	AR-1254-4							0.000	-1.00
30) L6	AR-1254-5							0.000	-1.00
31) L7	AR-1260-1							0.000	-1.00
32) L7	AR-1260-2							0.000	-1.00
33) L7	AR-1260-3							0.000	-1.00
34) L7	AR-1260-4							0.000	-1.00
35) L7	AR-1260-5							0.000	-1.00
36) L8	AR-1262-1							0.000	-1.00
37) L8	AR-1262-2							0.000	-1.00
38) L8	AR-1262-3							0.000	-1.00
39) L8	AR-1262-4							0.000	-1.00
40) L8	AR-1262-5							0.000	-1.00
41) L9	AR-1268-1							0.000	-1.00
42) L9	AR-1268-2							0.000	-1.00
43) L9	AR-1268-3							0.000	-1.00
44) L9	AR-1268-4							0.000	-1.00
45) L9	AR-1268-5							0.000	-1.00

 (#) = Out of Range