

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR060623CLP.M
 Title : GC EXTRACTABLES
 Last Update : Wed Jun 07 04:02:06 2023
 Response Via : Initial Calibration

Calibration Files

400 =PR061671.D 1600 =PR061673.D 800 =PR061672.D
 200 =PR061670.D 100 =PR061669.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	2.499	2.270	2.359	2.579	2.710	2.483	E6	7.03
2) SA	Decachlorobiphenyl	1.417	1.218	1.330	1.552	1.609	1.425	E6	11.20
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1	0.850	0.718	0.772	0.928	1.001	0.854	E5	13.35
27) L6	AR-1254-2	1.253	1.062	1.145	1.354	1.454	1.253	E5	12.53
28) L6	AR-1254-3	1.229	1.071	1.144	1.319	1.408	1.234	E5	10.87
29) L6	AR-1254-4	8.689	7.567	8.098	9.341	9.883	8.715	E4	10.66
30) L6	AR-1254-5	0.915	0.797	0.860	0.992	1.048	0.922	E5	10.88
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PR061671.D 1600 =PR061673.D 800 =PR061672.D
 200 =PR061670.D 100 =PR061669.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	3.186	2.858	2.927	3.286	3.465	3.145	E6	8.01
2) SA	Decachlorobiphenyl	3.778	3.453	3.560	4.047	4.243	3.816	E6	8.63
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR060623CLP.M
 Title : GC EXTRACTABLES
 Last Update : Wed Jun 07 04:02:06 2023
 Response Via : Initial Calibration

Calibration Files

400 =PR061671.D 1600 =PR061673.D 800 =PR061672.D
 200 =PR061670.D 100 =PR061669.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1	1.735	1.518	1.599	1.863	2.029	1.749 E5	11.70
27) L6	AR-1254-2	1.509	1.307	1.383	1.640	1.786	1.525 E5	12.67
28) L6	AR-1254-3	2.511	2.262	2.347	2.710	2.850	2.536 E5	9.66
29) L6	AR-1254-4	1.631	1.468	1.521	1.754	1.860	1.647 E5	9.84
30) L6	AR-1254-5	2.437	2.178	2.287	2.652	2.835	2.478 E5	10.78
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range