

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR060623CLP.M
 Title : GC EXTRACTABLES
 Last Update : Wed Jun 07 04:47:48 2023
 Response Via : Initial Calibration

Calibration Files

400 =PR061681.D 1600 =PR061683.D 800 =PR061682.D
 200 =PR061680.D 100 =PR061679.D

Compound			400	1600	800	200	100	Avg	%RSD	
1)	SA	Tetrachloro-m-xylene	2.392	2.291	2.391	2.576	2.692	2.468	E6	6.55
2)	SA	Decachlorobiphenyl	2.226	1.910	2.076	2.416	2.536	2.233	E6	11.31
3)	L1	AR-1016-1						0.000		-1.00
4)	L1	AR-1016-2						0.000		-1.00
5)	L1	AR-1016-3						0.000		-1.00
6)	L1	AR-1016-4						0.000		-1.00
7)	L1	AR-1016-5						0.000		-1.00
8)	L2	AR-1221-1						0.000		-1.00
9)	L2	AR-1221-2						0.000		-1.00
10)	L2	AR-1221-3						0.000		-1.00
11)	L3	AR-1232-1						0.000		-1.00
12)	L3	AR-1232-2						0.000		-1.00
13)	L3	AR-1232-3						0.000		-1.00
14)	L3	AR-1232-4						0.000		-1.00
15)	L3	AR-1232-5						0.000		-1.00
16)	L4	AR-1242-1						0.000		-1.00
17)	L4	AR-1242-2						0.000		-1.00
18)	L4	AR-1242-3						0.000		-1.00
19)	L4	AR-1242-4						0.000		-1.00
20)	L4	AR-1242-5						0.000		-1.00
21)	L5	AR-1248-1						0.000		-1.00
22)	L5	AR-1248-2						0.000		-1.00
23)	L5	AR-1248-3						0.000		-1.00
24)	L5	AR-1248-4						0.000		-1.00
25)	L5	AR-1248-5						0.000		-1.00
26)	L6	AR-1254-1						0.000		-1.00
27)	L6	AR-1254-2						0.000		-1.00
28)	L6	AR-1254-3						0.000		-1.00
29)	L6	AR-1254-4						0.000		-1.00
30)	L6	AR-1254-5						0.000		-1.00
31)	L7	AR-1260-1						0.000		-1.00
32)	L7	AR-1260-2						0.000		-1.00
33)	L7	AR-1260-3						0.000		-1.00
34)	L7	AR-1260-4						0.000		-1.00
35)	L7	AR-1260-5						0.000		-1.00
36)	L8	AR-1262-1						0.000		-1.00
37)	L8	AR-1262-2						0.000		-1.00
38)	L8	AR-1262-3						0.000		-1.00
39)	L8	AR-1262-4						0.000		-1.00
40)	L8	AR-1262-5						0.000		-1.00
41)	L9	AR-1268-1	1.644	1.452	1.560	1.824	1.914	1.679	E5	11.28
42)	L9	AR-1268-2	1.506	1.324	1.430	1.707	1.851	1.563	E5	13.63
43)	L9	AR-1268-3	1.290	1.141	1.232	1.420	1.464	1.310	E5	10.17
44)	L9	AR-1268-4	5.377	4.681	5.184	5.865	6.135	5.448	E4	10.50
45)	L9	AR-1268-5	3.855	3.515	3.760	4.205	4.398	3.946	E5	8.95

Signal #2 Calibration Files

400 =PR061681.D 1600 =PR061683.D 800 =PR061682.D
 200 =PR061680.D 100 =PR061679.D

Compound			400	1600	800	200	100	Avg	%RSD	
1)	SA	Tetrachloro-m-xylene	3.036	2.870	3.012	3.330	3.481	3.146	E6	8.00
2)	SA	Decachlorobiphenyl	5.901	5.437	5.750	6.485	6.727	6.060	E6	8.79
3)	L1	AR-1016-1						0.000		-1.00
4)	L1	AR-1016-2						0.000		-1.00
5)	L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR060623CLP.M
 Title : GC EXTRACTABLES
 Last Update : Wed Jun 07 04:47:48 2023
 Response Via : Initial Calibration

Calibration Files

400 =PR061681.D 1600 =PR061683.D 800 =PR061682.D
 200 =PR061680.D 100 =PR061679.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1	4.312	3.847	4.104	4.727	4.893	4.377 E5	9.88
42) L9	AR-1268-2	3.929	3.659	3.874	4.302	4.421	4.037 E5	7.82
43) L9	AR-1268-3	3.430	3.203	3.365	3.721	3.809	3.506 E5	7.21
44) L9	AR-1268-4	1.536	1.378	1.481	1.713	1.750	1.571 E5	10.01
45) L9	AR-1268-5	1.112	1.061	1.108	1.191	1.202	1.135 E6	5.30

(#) = Out of Range