

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR060721\
 Data File : PR050688.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Jun 2021 13:17
 Operator : DD\AJ
 Sample : AR1221ICC500
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR1221ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 08 02:23:04 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR060721.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jun 08 02:22:32 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.632	3.815	320.7E6	325.2E6	50.000	50.000
2) SA Decachlor...	10.506	8.843	363.0E6	381.0E6	50.000	50.000
Target Compounds						
8) L2 AR-1221-1	4.845	4.027	34654997	33562282	500.000	500.000
9) L2 AR-1221-2	4.933	4.112	26200651	25762953	500.000	500.000
10) L2 AR-1221-3	5.012	4.189	82136045	81466436	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR060721\
Data File : PR050688.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Jun 2021 13:17
Operator : DD\AJ
Sample : AR1221ICC500
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR1221ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 08 02:23:04 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR060721.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jun 08 02:22:32 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.632 min
Delta R.T.: 0.000 min
Response: 320677318
Conc: 50.00 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1221ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.815 min
Delta R.T.: 0.000 min
Response: 325177389
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.506 min
Delta R.T.: 0.000 min
Response: 363034367
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.843 min
Delta R.T.: 0.000 min
Response: 381004186
Conc: 50.00 ng/ml



#8 AR-1221-1


R.T.: 4.845 min
Delta R.T.: 0.000 min
Response: 34654997
Conc: 500.00 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1221ICC500




#8 AR-1221-1

R.T.: 4.027 min
Delta R.T.: 0.000 min
Response: 33562282
Conc: 500.00 ng/ml




#9 AR-1221-2

R.T.: 4.933 min
Delta R.T.: 0.000 min
Response: 26200651
Conc: 500.00 ng/ml



#9 AR-1221-2


R.T.: 4.112 min
Delta R.T.: 0.000 min
Response: 25762953
Conc: 500.00 ng/ml



#10 AR-1221-3

R.T.: 5.012 min
Delta R.T.: 0.000 min
Response: 82136045
Conc: 500.00 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1221ICC500



#10 AR-1221-3

R.T.: 4.189 min
Delta R.T.: 0.000 min
Response: 81466436
Conc: 500.00 ng/ml