

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR060722\
 Data File : PR054748.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Jun 2022 23:50
 Operator : AJ\MA
 Sample : N3194-04MS
 Misc :
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 TP-3MS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 08 01:42:48 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR053122.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 01 05:23:25 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.650	2.924	88088045	40043211	21.822	21.596
2) SA Decachlor...	9.534	8.145	25957950	26401414	15.883	15.785
Target Compounds						
3) L1 AR-1016-1	4.866	4.031	61932198	34354498	485.923	521.120
4) L1 AR-1016-2	4.889	4.049	87793241	47079170	489.659	526.059
5) L1 AR-1016-3	4.953	4.225	53396107	25865941	488.199	532.858
6) L1 AR-1016-4	5.055	4.276	43181822	19525958	491.316	517.026
7) L1 AR-1016-5	5.370	4.492	37439458	25276550	469.795	519.845
31) L7 AR-1260-1	6.581	5.576	47264883	47810150	459.494	510.532
32) L7 AR-1260-2	6.862	5.780	52337342	56963652	439.498	493.642
33) L7 AR-1260-3	7.253	5.937	33869269	53788948	442.296	449.076
34) L7 AR-1260-4	7.494	6.441	40968152	38513951	454.435	467.874
35) L7 AR-1260-5	7.832	6.704	70632217	90612695	422.332	465.637



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR060722\
Data File : PR054748.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Jun 2022 23:50
Operator : AJ\MA
Sample : N3194-04MS
Misc :
ALS Vial : 47 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
TP-3MS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 08 01:42:48 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR053122.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 01 05:23:25 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.650 min
Delta R.T.: 0.000 min
Response: 88088045
Conc: 21.82 ng/ml

Instrument :
ECD_R
ClientSampleId :
TP-3MS




#1 Tetrachloro-m-xylene

R.T.: 2.924 min
Delta R.T.: 0.000 min
Response: 40043211
Conc: 21.60 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.534 min
Delta R.T.: 0.000 min
Response: 25957950
Conc: 15.88 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.145 min
Delta R.T.: -0.013 min
Response: 26401414
Conc: 15.78 ng/ml



#3 AR-1016-1


R.T.: 4.866 min
Delta R.T.: -0.002 min
Response: 61932198
Conc: 485.92 ng/ml

Instrument :
ECD_R
ClientSampleId :
TP-3MS




#3 AR-1016-1

R.T.: 4.031 min
Delta R.T.: -0.003 min
Response: 34354498
Conc: 521.12 ng/ml




#4 AR-1016-2

R.T.: 4.889 min
Delta R.T.: -0.002 min
Response: 87793241
Conc: 489.66 ng/ml



#4 AR-1016-2


R.T.: 4.049 min
Delta R.T.: -0.003 min
Response: 47079170
Conc: 526.06 ng/ml



#5 AR-1016-3


R.T.: 4.953 min
Delta R.T.: -0.001 min
Response: 53396107
Conc: 488.20 ng/ml

Instrument :
ECD_R
ClientSampleId :
TP-3MS




#5 AR-1016-3

R.T.: 4.225 min
Delta R.T.: -0.004 min
Response: 25865941
Conc: 532.86 ng/ml




#6 AR-1016-4

R.T.: 5.055 min
Delta R.T.: -0.002 min
Response: 43181822
Conc: 491.32 ng/ml



#6 AR-1016-4


R.T.: 4.276 min
Delta R.T.: -0.004 min
Response: 19525958
Conc: 517.03 ng/ml



#7 AR-1016-5


R.T.: 5.370 min
Delta R.T.: -0.002 min
Response: 37439458
Conc: 469.80 ng/ml

Instrument :
ECD_R
ClientSampleId :
TP-3MS




#7 AR-1016-5

R.T.: 4.492 min
Delta R.T.: -0.003 min
Response: 25276550
Conc: 519.84 ng/ml




#31 AR-1260-1

R.T.: 6.581 min
Delta R.T.: -0.002 min
Response: 47264883
Conc: 459.49 ng/ml



#31 AR-1260-1


R.T.: 5.576 min
Delta R.T.: -0.004 min
Response: 47810150
Conc: 510.53 ng/ml



#32 AR-1260-2


R.T.: 6.862 min
Delta R.T.: -0.003 min
Response: 52337342
Conc: 439.50 ng/ml

Instrument :
ECD_R
ClientSampleId :
TP-3MS




#32 AR-1260-2

R.T.: 5.780 min
Delta R.T.: -0.004 min
Response: 56963652
Conc: 493.64 ng/ml




#33 AR-1260-3

R.T.: 7.253 min
Delta R.T.: -0.003 min
Response: 33869269
Conc: 442.30 ng/ml



#33 AR-1260-3


R.T.: 5.937 min
Delta R.T.: -0.006 min
Response: 53788948
Conc: 449.08 ng/ml



#34 AR-1260-4


R.T.: 7.494 min
Delta R.T.: -0.002 min
Response: 40968152
Conc: 454.44 ng/ml

Instrument :
ECD_R
ClientSampleId :
TP-3MS




#34 AR-1260-4

R.T.: 6.441 min
Delta R.T.: -0.006 min
Response: 38513951
Conc: 467.87 ng/ml



#35 AR-1260-5

R.T.: 7.832 min
Delta R.T.: -0.003 min
Response: 70632217
Conc: 422.33 ng/ml



#35 AR-1260-5

R.T.: 6.704 min
Delta R.T.: -0.007 min
Response: 90612695
Conc: 465.64 ng/ml