

Data Path : Z:\PESTPCBSRV\HPCHEM1\ECD_R\Data\PR061018\
 Data File : PR028773.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Jun 2018 19:21
 Operator : UA/SM
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 I.BLK

Manual Integrations
 APPROVED

UGO
 6/6/2018 7:54:45 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 06 19:51:04 2018
 Quant Method : \\TERASTORAGE\Terastorage\PESTPCBSRV\HPCHEM1\ECD_R\Method\PR061018.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 06 14:28:54 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.584	3.727	538.6E6	895.3E6	21.822m	18.362
2) SA Decachlor...	10.370	8.685	384.4E6	1086.8E6	18.074	19.061

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\PESTPCBSRV\HPCHEM1\ECD_R\Data\PR061018\
Data File : PR028773.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Jun 2018 19:21
Operator : UA/SM
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
I.BLK

Manual Integrations
APPROVED

UGO
6/6/2018 7:54:45 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 06 19:51:04 2018
Quant Method : \\TERASTORAGE\Terastorage\PESTPCBSRV\HPCHEM1\ECD_R\Method\PR061018.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 06 14:28:54 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

