

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR061019\
 Data File : PR038616.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Jun 2019 12:11
 Operator : SM\MA
 Sample : AR1248ICC500
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR1248ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 10 14:50:25 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR061019.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Jun 10 14:40:36 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.772	4.619	99284489	330.6E6	50.000	50.000
2)	SA Decachlor...	8.730	10.370	101.9E6	273.2E6	50.000	50.000
Target Compounds							
21)	L5 AR-1248-1	4.844	5.775	21674113	67424466	500.000	500.000
22)	L5 AR-1248-2	5.076	6.043	29182277	88068910	500.000	500.000
23)	L5 AR-1248-3	5.117	6.248	29820794	100.4E6	500.000	500.000
24)	L5 AR-1248-4	5.287	6.646	36546026	117.2E6	500.000	500.000
25)	L5 AR-1248-5	5.673	6.685	38198041	110.7E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR061019\
Data File : PR038616.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Jun 2019 12:11
Operator : SM\MA
Sample : AR1248ICC500
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR1248ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 10 14:50:25 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR061019.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jun 10 14:40:36 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

