

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR061321\
 Data File : PR050796.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Jun 2021 09:27
 Operator : DD\AJ
 Sample : M2558-10
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 X3556

Manual Integrations
 APPROVED

Ankita
 6/14/2021 2:36:04 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 14 07:49:56 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR060121CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 02 05:07:41 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.631	3.813	144.9E6	139.2E6	22.700	21.408
2) SA Decachlor...	10.499	8.836	315.4E6	318.4E6	43.413	41.054
Target Compounds						
21) L5 AR-1248-1	5.812	4.898	59818724	53106208	375.337	358.038
22) L5 AR-1248-2	6.085	5.134	84712160	72497775	366.148	345.054
23) L5 AR-1248-3	6.290	5.176	94221719	76515063	357.628m	343.053
24) L5 AR-1248-4	6.697	5.347	113.4E6	93389740	355.557	338.760
25) L5 AR-1248-5	6.736	5.739	111.1E6	96394329	362.588	328.150

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR061321\
Data File : PR050796.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Jun 2021 09:27
Operator : DD\AJ
Sample : M2558-10
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampled :
X3556

Manual Integrations
APPROVED

Ankita
6/14/2021 2:36:04 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 14 07:49:56 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR060121CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 02 05:07:41 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

