

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR061521\
Data File : PR050854.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Jun 2021 13:11
Operator : DD\AJ
Sample : AR1660CCC400
Misc :
ALS Vial : 3 Sample Multiplier: 1


Instrument :
ECD_R
LabSampleId :
AR1660CCC400

Manual Integrations
APPROVED

Ankita
6/18/2021 12:10:53 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 16 08:02:07 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR060121CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 02 05:07:41 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(31) AR-1260-1 (L7)
R.T. Response Conc
7.42 168915057 475.23
7.68 139460174 322.67
8.04 102678561 317.58
8.27 118340597 308.04
8.61 253398156 316.26

(31) AR-1260-1 #2 (L7)
R.T. Response Conc
6.38 136425289 339.14
6.56 164687941 337.46
6.72 156533683 334.92
7.19 129135284 323.62
7.43 324144561 332.88

(+) = Expected Retention Time

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR061521\
Data File : PR050854.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Jun 2021 13:11
Operator : DD\AJ
Sample : AR1660CCC400
Misc :
ALS Vial : 3 Sample Multiplier: 1


Instrument :
ECD_R
LabSampleId :
AR1660CCC400

Manual Integrations
APPROVED

Ankita
6/18/2021 12:10:53 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 16 08:02:07 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR060121CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 02 05:07:41 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(31) AR-1260-1 #2 (L7)
R.T. Response Conc
7.42 141494191 398.08
7.68 139460174 322.67
8.04 102678561 317.58
8.27 118340597 308.04
8.61 253398156 316.26

(31) AR-1260-1 #2 (L7)
R.T. Response Conc
6.38 136425289 339.14
6.56 164687941 337.46
6.72 156533683 334.92
7.19 129135284 323.62
7.43 324144561 332.88

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR061521\
 Data File : PR050854.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Jun 2021 13:11
 Operator : DD\AJ
 Sample : AR1660CCC400
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_R
 LabSampleId :
 AR1660CCC400

Manual Integrations
 APPROVED

Ankita
 6/18/2021 12:10:53 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 16 08:02:07 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR060121CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 02 05:07:41 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.633	3.814	146.6E6	134.8E6	22.972	20.727
2) SA Decachlor...	10.509	8.841	253.5E6	233.5E6	34.887	30.116
Target Compounds						
3) L1 AR-1016-1	5.815	4.899	121.6E6	101.4E6	446.723	421.444
4) L1 AR-1016-2	5.838	4.918	169.0E6	147.7E6	439.626	402.473
5) L1 AR-1016-3	5.900	5.094	113.1E6	76088015	493.178	399.920
6) L1 AR-1016-4	6.001	5.135	78024302	57310716	405.120	370.451
7) L1 AR-1016-5	6.295	5.348	75381499	78293200	388.676	382.621
31) L7 AR-1260-1	7.420	6.378	141.5E6	136.4E6	398.083m	339.141
32) L7 AR-1260-2	7.678	6.565	139.5E6	164.7E6	322.671	337.455
33) L7 AR-1260-3	8.037	6.719	102.7E6	156.5E6	317.580	334.922
34) L7 AR-1260-4	8.273	7.190	118.3E6	129.1E6	308.035	323.616
35) L7 AR-1260-5	8.610	7.429	253.4E6	324.1E6	316.255	332.875

AJ
 06/18/21

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.