

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR061724\
 Data File : PR067396.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Jun 2024 15:02
 Operator : AJ\MA
 Sample : AR1660ICC800
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR16604229

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/18/2024
 Supervised By :Ankita Jodhani 06/18/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 18 02:05:25 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR061724CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jun 18 02:05:17 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.648	2.884	182.8E6	188.0E6	38.383	37.520
2) SA Decachlor...	9.528	8.107	108.6E6	142.4E6	74.455	78.589
Target Compounds						
3) L1 AR-1016-1	4.867	3.991	99119756	126.1E6	788.108m	737.186
4) L1 AR-1016-2	4.890	4.009	156.3E6	183.8E6	728.347m	742.266
5) L1 AR-1016-3	4.954	4.186	94036879	95006298	706.064	739.363
6) L1 AR-1016-4	5.057	4.238	79781528	72657550	749.826	726.994
7) L1 AR-1016-5	5.372	4.453	79198302	92459639	759.796	730.368
31) L7 AR-1260-1	6.581	5.537	145.1E6	213.2E6	726.661	729.644
32) L7 AR-1260-2	6.863	5.742	160.8E6	267.3E6	731.020	727.487
33) L7 AR-1260-3	7.252	5.899	122.2E6	249.0E6	726.731	771.826
34) L7 AR-1260-4	7.494	6.402	135.8E6	204.6E6	742.661	734.551
35) L7 AR-1260-5	7.832	6.667	252.0E6	451.4E6	761.785	794.524

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR061724\
Data File : PR067396.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Jun 2024 15:02
Operator : AJ\MA
Sample : AR1660ICC800
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR16604229

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/18/2024
Supervised By :Ankita Jodhani 06/18/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 18 02:05:25 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR061724CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jun 18 02:05:17 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

