

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR061819\
 Data File : PR038771.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Jun 2019 23:18
 Operator : SM\MA
 Sample : K3362-01
 Misc :
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 A41T7

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 19 04:50:17 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR061819CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 19 04:41:49 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.782	4.631	7730330	26333066	11.585	10.202
2)	SA Decachlor...	8.731	10.379	44625258	124.2E6	21.237	19.650
Target Compounds							
20)	L4 AR-1242-5	0.000	6.629f	0	3045017	N.D.	31.657 #
24)	L5 AR-1248-4	0.000	6.629	0	3045017	N.D.	20.679 #
25)	L5 AR-1248-5	0.000	6.629f	0	3045017	N.D.	21.454 #
26)	L6 AR-1254-1	0.000	6.629	0	3045017	N.D.	19.406 #
27)	L6 AR-1254-2	0.000	6.847	0	404597	N.D.	1.584 #
28)	L6 AR-1254-3	0.000	7.221	0	3883362	N.D.	12.721 #
29)	L6 AR-1254-4	0.000	7.488	0	3663692	N.D.	15.188 #
32)	L7 AR-1260-2	6.484	7.622	747458	459806	7.335	1.751 #
38)	L8 AR-1262-3	7.622	0.000	553918	0	4.938	N.D. #
40)	L8 AR-1262-5	8.190	0.000	1249651	0	10.620	N.D. #
41)	L9 AR-1268-1	7.622	0.000	553918	0	1.620	N.D. #
43)	L9 AR-1268-3	0.000	9.174	0	887824	N.D.	1.348 #
44)	L9 AR-1268-4	8.190	0.000	1249651	0	9.086	N.D. #
45)	L9 AR-1268-5	8.476	10.037	1736014	1070603	1.871	0.434 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR061819\
Data File : PR038771.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Jun 2019 23:18
Operator : SM\MA
Sample : K3362-01
Misc :
ALS Vial : 49 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampled :
A41T7

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 19 04:50:17 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR061819CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 19 04:41:49 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#20 AR-1242-5


R.T.: 0.000 min
Exp R.T.: 5.639 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
A41T7




#20 AR-1242-5

R.T.: 6.629 min
Delta R.T.: -0.065 min
Response: 3045017
Conc: 31.66 ng/ml




#24 AR-1248-4

R.T.: 0.000 min
Exp R.T.: 5.294 min
Response: 0
Conc: N.D.



#24 AR-1248-4


R.T.: 6.629 min
Delta R.T.: -0.026 min
Response: 3045017
Conc: 20.68 ng/ml



#25 AR-1248-5


R.T.: 0.000 min
Exp R.T.: 5.680 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
A41T7




#25 AR-1248-5

R.T.: 6.629 min
Delta R.T.: -0.065 min
Response: 3045017
Conc: 21.45 ng/ml




#26 AR-1254-1

R.T.: 0.000 min
Exp R.T.: 5.639 min
Response: 0
Conc: N.D.



#26 AR-1254-1


R.T.: 6.629 min
Delta R.T.: 0.000 min
Response: 3045017
Conc: 19.41 ng/ml



#27 AR-1254-2


R.T.: 0.000 min
Exp R.T.: 5.783 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
A41T7




#27 AR-1254-2

R.T.: 6.847 min
Delta R.T.: 0.004 min
Response: 404597
Conc: 1.58 ng/ml




#28 AR-1254-3

R.T.: 0.000 min
Exp R.T.: 6.181 min
Response: 0
Conc: N.D.



#28 AR-1254-3


R.T.: 7.221 min
Delta R.T.: 0.013 min
Response: 3883362
Conc: 12.72 ng/ml



#29 AR-1254-4


R.T.: 0.000 min
Exp R.T.: 6.405 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
A41T7




#29 AR-1254-4

R.T.: 7.488 min
Delta R.T.: -0.002 min
Response: 3663692
Conc: 15.19 ng/ml




#32 AR-1260-2

R.T.: 6.484 min
Delta R.T.: -0.011 min
Response: 747458
Conc: 7.33 ng/ml



#32 AR-1260-2


R.T.: 7.622 min
Delta R.T.: 0.000 min
Response: 459806
Conc: 1.75 ng/ml



#38 AR-1262-3


R.T.: 7.622 min
Delta R.T.: -0.008 min
Response: 553918
Conc: 4.94 ng/ml

Instrument :
ECD_R
ClientSampleId :
A41T7




#38 AR-1262-3

R.T.: 0.000 min
Exp R.T. : 8.850 min
Response: 0
Conc: N.D.




#40 AR-1262-5

R.T.: 8.190 min
Delta R.T.: 0.006 min
Response: 1249651
Conc: 10.62 ng/ml



#40 AR-1262-5


R.T.: 0.000 min
Exp R.T. : 9.614 min
Response: 0
Conc: N.D.



#41 AR-1268-1


R.T.: 7.622 min
Delta R.T.: -0.008 min
Response: 553918
Conc: 1.62 ng/ml

Instrument :
ECD_R
ClientSampleId :
A41T7




#41 AR-1268-1

R.T.: 0.000 min
Exp R.T. : 8.849 min
Response: 0
Conc: N.D.




#43 AR-1268-3

R.T.: 0.000 min
Exp R.T. : 7.898 min
Response: 0
Conc: N.D.



#43 AR-1268-3


R.T.: 9.174 min
Delta R.T.: -0.005 min
Response: 887824
Conc: 1.35 ng/ml



#44 AR-1268-4


R.T.: 8.190 min
Delta R.T.: 0.006 min
Response: 1249651
Conc: 9.09 ng/ml

Instrument :
ECD_R
ClientSampleId :
A41T7




#44 AR-1268-4

R.T.: 0.000 min
Exp R.T. : 9.615 min
Response: 0
Conc: N.D.



#45 AR-1268-5

R.T.: 8.476 min
Delta R.T.: -0.002 min
Response: 1736014
Conc: 1.87 ng/ml



#45 AR-1268-5

R.T.: 10.037 min
Delta R.T.: 0.000 min
Response: 1070603
Conc: 0.43 ng/ml