

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR061819\
 Data File : PR038778.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Jun 2019 01:00
 Operator : SM\MA
 Sample : K3362-17
 Misc :
 ALS Vial : 56 Sample Multiplier: 1


Instrument :
 ECD_R
 ClientSampleId :
 A4210

Manual Integrations
 APPROVED

Ankita
 6/20/2019 10:00:44 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 19 10:17:12 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR061819CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 19 04:41:49 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(16) AR-1242-1 (L4)
 R.T. Response Conc
 4.87 3322377 124.41
 4.87 3322377 76.71
 5.04 1200030 56.35
 5.12 2411146 106.45
 5.64 1152417 32.46

(16) AR-1242-1 #2 (L4)
 R.T. Response Conc
 5.79 2817384 30.31
 5.81 7575257 57.02
 5.87 809884 9.67
 5.97 3556930 51.37
 6.69 4225562 43.93

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR061819\
Data File : PR038778.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Jun 2019 01:00
Operator : SM\MA
Sample : K3362-17
Misc :
ALS Vial : 56 Sample Multiplier: 1


Instrument :
ECD_R
ClientSampleId :
A4210

Manual Integrations
APPROVED

Ankita
6/20/2019 10:00:44 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 19 10:17:12 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR061819CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 19 04:41:49 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(16) AR-1242-1 (L4)
R.T. Response Conc
4.86 825427 30.91
4.87 2284792 52.75
0.00 0 0.00
5.12 2411146 106.45
5.64 1152417 32.46

(16) AR-1242-1 #2 (L4)
R.T. Response Conc
5.79 2817384 30.31
5.81 7575257 57.02
0.00 0 0.00
5.97 3556930 51.37
6.69 4225562 43.93