

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR061923\
 Data File : PR061869.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Jun 2023 13:18
 Operator : YP\AJ
 Sample : AR1268ICC500
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR1268ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 19 13:26:33 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR061923.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Jun 19 13:26:18 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.691	2.962	111.2E6	144.8E6	50.000	50.000
2) SA Decachlor...	9.625	8.249	117.0E6	299.8E6	50.000	50.000
Target Compounds						
41) L9 AR-1268-1	8.202	7.098	98870750	257.2E6	500.000	500.000
42) L9 AR-1268-2	8.294	7.167	91099842	237.9E6	500.000	500.000
43) L9 AR-1268-3	8.514	7.390	78224436	204.8E6	500.000	500.000
44) L9 AR-1268-4	8.921	7.704	32175660	87878360	500.000	500.000
45) L9 AR-1268-5	9.310	8.000	230.0E6	644.1E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR061923\
Data File : PR061869.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Jun 2023 13:18
Operator : YP\AJ
Sample : AR1268ICC500
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR1268ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 19 13:26:33 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR061923.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jun 19 13:26:18 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

