

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD061924\
 Data File : PD083746.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Jun 2024 15:24
 Operator : AR\AJ
 Sample : P2899-08
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 C0AB0

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 19 22:34:06 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060724CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Jun 07 04:03:14 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.561	2.772	14060318	67938059	10.772	12.463
27)	SA Decachlor...	9.114	7.896	26480047	85464624	13.618	14.075
Target Compounds							
4)	MA Heptachlor	0.000	3.990f	0	33549453	N.D.	4.285
7)	B delta-BHC	4.793	0.000	1206629	0	0.459	N.D. #
12)	B 4,4'-DDE	6.242	0.000	3365212	0	1.622	N.D. #
21)	B Endrin ke...	7.700	0.000	853959	0	0.458	N.D. #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD061924\
Data File : PD083746.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Jun 2024 15:24
Operator : AR\AJ
Sample : P2899-08
Misc :
ALS Vial : 13 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 19 22:34:06 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD060724CLP.M
Quant Title : GC Extractables
QLast Update : Fri Jun 07 04:03:14 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.561 min
Delta R.T.: -0.002 min
Response: 14060318
Conc: 10.77 ng/ml

Instrument :
ECD_R
ClientSampleId :
C0AB0




#1 Tetrachloro-m-xylene

R.T.: 2.772 min
Delta R.T.: -0.001 min
Response: 67938059
Conc: 12.46 ng/ml




#4 Heptachlor

R.T.: 0.000 min
Exp R.T. : 4.947 min
Response: 0
Conc: N.D.



#4 Heptachlor


R.T.: 3.990 min
Delta R.T.: 0.050 min
Response: 33549453
Conc: 4.29 ng/ml



#7 delta-BHC


R.T.: 4.793 min
Delta R.T.: -0.009 min
Response: 1206629
Conc: 0.46 ng/ml

Instrument :
ECD_R
ClientSampleId :
C0AB0




#7 delta-BHC

R.T.: 0.000 min
Exp R.T. : 4.128 min
Response: 0
Conc: N.D.




#12 4,4'-DDE

R.T.: 6.242 min
Delta R.T.: 0.015 min
Response: 3365212
Conc: 1.62 ng/ml



#12 4,4'-DDE


R.T.: 0.000 min
Exp R.T. : 5.223 min
Response: 0
Conc: N.D.



#21 Endrin ketone


R.T.: 7.700 min
Delta R.T.: 0.014 min
Response: 853959
Conc: 0.46 ng/ml

Instrument :
ECD_R
ClientSampleId :
C0AB0




#21 Endrin ketone

R.T.: 0.000 min
Exp R.T. : 6.830 min
Response: 0
Conc: N.D.



#27 Decachlorobiphenyl

R.T.: 9.114 min
Delta R.T.: -0.007 min
Response: 26480047
Conc: 13.62 ng/ml



#27 Decachlorobiphenyl

R.T.: 7.896 min
Delta R.T.: -0.005 min
Response: 85464624
Conc: 14.07 ng/ml