

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR062320CLP.M
 Title : GC EXTRACTABLES
 Last Update : Tue Jun 23 02:09:57 2020
 Response Via : Initial Calibration

Calibration Files

400 =PR045925.D 1600 =PR045927.D 800 =PR045926.D
 200 =PR045924.D 100 =PR045923.D

Compound		400	1600	800	200	100	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	1.686	1.693	1.660	1.672	1.765	1.695	E6	2.42
2) SA	Decachlorobiphenyl	1.313	1.234	1.273	1.382	1.380	1.317	E6	4.95
3) L1	AR-1016-1	5.232	5.435	4.839	5.650	6.079	5.447	E4	8.50
4) L1	AR-1016-2	8.189	7.866	7.935	8.526	8.965	8.296	E4	5.48
5) L1	AR-1016-3	5.053	4.582	4.686	5.270	5.414	5.001	E4	7.21
6) L1	AR-1016-4	4.126	3.843	3.927	4.348	4.520	4.153	E4	6.81
7) L1	AR-1016-5	4.008	3.640	3.770	4.008	4.256	3.936	E4	6.07
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1	7.155	6.663	6.870	7.467	7.776	7.187	E4	6.23
32) L7	AR-1260-2	8.622	8.147	8.356	8.981	9.181	8.657	E4	4.94
33) L7	AR-1260-3	6.561	6.248	6.413	6.850	6.826	6.579	E4	3.96
34) L7	AR-1260-4	7.710	7.381	7.554	7.993	8.059	7.739	E4	3.71
35) L7	AR-1260-5	1.579	1.572	1.576	1.617	1.635	1.596	E5	1.78
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PR045925.D 1600 =PR045927.D 800 =PR045926.D
 200 =PR045924.D 100 =PR045923.D

Compound		400	1600	800	200	100	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	1.144	1.139	1.126	1.128	1.207	1.149	E7	2.91
2) SA	Decachlorobiphenyl	0.994	0.940	0.970	1.033	1.050	0.997	E7	4.52
3) L1	AR-1016-1	4.280	3.900	4.137	4.395	4.774	4.297	E5	7.54
4) L1	AR-1016-2	4.672	5.486	4.273	5.032	5.622	5.017	E5	11.18
5) L1	AR-1016-3	2.389	2.489	2.477	2.456	2.646	2.491	E5	3.80

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR062320CLP.M
 Title : GC EXTRACTABLES
 Last Update : Tue Jun 23 02:09:57 2020
 Response Via : Initial Calibration

Calibration Files

400 =PR045925.D 1600 =PR045927.D 800 =PR045926.D
 200 =PR045924.D 100 =PR045923.D

Compound			400	1600	800	200	100	Avg	%RSD

6)	L1	AR-1016-4	1.999	1.942	1.991	2.140	2.578	2.130 E5	12.25
7)	L1	AR-1016-5	3.057	2.884	2.951	3.160	3.435	3.098 E5	6.98
8)	L2	AR-1221-1						0.000	-1.00
9)	L2	AR-1221-2						0.000	-1.00
10)	L2	AR-1221-3						0.000	-1.00
11)	L3	AR-1232-1						0.000	-1.00
12)	L3	AR-1232-2						0.000	-1.00
13)	L3	AR-1232-3						0.000	-1.00
14)	L3	AR-1232-4						0.000	-1.00
15)	L3	AR-1232-5						0.000	-1.00
16)	L4	AR-1242-1						0.000	-1.00
17)	L4	AR-1242-2						0.000	-1.00
18)	L4	AR-1242-3						0.000	-1.00
19)	L4	AR-1242-4						0.000	-1.00
20)	L4	AR-1242-5						0.000	-1.00
21)	L5	AR-1248-1						0.000	-1.00
22)	L5	AR-1248-2						0.000	-1.00
23)	L5	AR-1248-3						0.000	-1.00
24)	L5	AR-1248-4						0.000	-1.00
25)	L5	AR-1248-5						0.000	-1.00
26)	L6	AR-1254-1						0.000	-1.00
27)	L6	AR-1254-2						0.000	-1.00
28)	L6	AR-1254-3						0.000	-1.00
29)	L6	AR-1254-4						0.000	-1.00
30)	L6	AR-1254-5						0.000	-1.00
31)	L7	AR-1260-1	5.520	5.286	5.429	5.731	6.009	5.595 E5	5.05
32)	L7	AR-1260-2	6.727	6.489	6.644	6.964	7.129	6.791 E5	3.76
33)	L7	AR-1260-3	6.191	6.088	6.178	6.315	6.406	6.235 E5	2.01
34)	L7	AR-1260-4	5.259	5.175	5.248	5.378	5.398	5.292 E5	1.78
35)	L7	AR-1260-5	1.338	1.329	1.343	1.352	1.346	1.342 E6	0.66
36)	L8	AR-1262-1						0.000	-1.00
37)	L8	AR-1262-2						0.000	-1.00
38)	L8	AR-1262-3						0.000	-1.00
39)	L8	AR-1262-4						0.000	-1.00
40)	L8	AR-1262-5						0.000	-1.00
41)	L9	AR-1268-1						0.000	-1.00
42)	L9	AR-1268-2						0.000	-1.00
43)	L9	AR-1268-3						0.000	-1.00
44)	L9	AR-1268-4						0.000	-1.00
45)	L9	AR-1268-5						0.000	-1.00

 (#) = Out of Range