

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR062418\
 Data File : PR029264.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Jun 2018 12:49
 Operator : UA/SM
 Sample : AR1254ICC500
 Misc :
 ALS Vial : 110 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR1254ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 19 13:29:06 2018
 Quant Method : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_R\Method\PR062418.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jun 19 13:28:58 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.576	3.721	1198.1E6	2067.8E6	49.038	49.422
2) SA Decachlor...	10.352	8.666	1198.4E6	1017.1E6	44.620	46.015
Target Compounds						
26) L6 AR-1254-1	6.006	5.024	282.7E6	613.6E6	500.000	500.000
27) L6 AR-1254-2	6.802	5.726	821.0E6	1450.8E6	500.000	500.000
28) L6 AR-1254-3	7.168	6.126	889.2E6	2423.8E6	500.000	500.000
29) L6 AR-1254-4	7.451	6.351	685.9E6	1868.2E6	500.000	500.000
30) L6 AR-1254-5	7.869	6.763	717.7E6	1905.6E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR062418\
Data File : PR029264.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Jun 2018 12:49
Operator : UA/SM
Sample : AR1254ICC500
Misc :
ALS Vial : 110 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR1254ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 19 13:29:06 2018
Quant Method : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_R\Method\PR062418.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jun 19 13:28:58 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.576 min
Delta R.T.: 0.000 min
Response: 1198080737
Conc: 49.04 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1254ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.721 min
Delta R.T.: 0.000 min
Response: 2067779356
Conc: 49.42 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.352 min
Delta R.T.: 0.000 min
Response: 1198374831
Conc: 44.62 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.666 min
Delta R.T.: 0.000 min
Response: 1017098142
Conc: 46.01 ng/ml



#26 AR-1254-1


R.T.: 6.006 min
Delta R.T.: 0.000 min
Response: 282677186
Conc: 500.00 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1254ICC500




#26 AR-1254-1

R.T.: 5.024 min
Delta R.T.: -0.008 min
Response: 613577431
Conc: 500.00 ng/ml




#27 AR-1254-2

R.T.: 6.802 min
Delta R.T.: -0.002 min
Response: 820982314
Conc: 500.00 ng/ml



#27 AR-1254-2


R.T.: 5.726 min
Delta R.T.: -0.009 min
Response: 1450775027
Conc: 500.00 ng/ml



#28 AR-1254-3


R.T.: 7.168 min
Delta R.T.: -0.002 min
Response: 889178933
Conc: 500.00 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1254ICC500




#28 AR-1254-3

R.T.: 6.126 min
Delta R.T.: -0.010 min
Response: 2423794194
Conc: 500.00 ng/ml




#29 AR-1254-4

R.T.: 7.451 min
Delta R.T.: -0.002 min
Response: 685931356
Conc: 500.00 ng/ml



#29 AR-1254-4


R.T.: 6.351 min
Delta R.T.: -0.010 min
Response: 1868224125
Conc: 500.00 ng/ml



#30 AR-1254-5

R.T.: 7.869 min
Delta R.T.: -0.002 min
Response: 717680823
Conc: 500.00 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1254ICC500



#30 AR-1254-5

R.T.: 6.763 min
Delta R.T.: -0.011 min
Response: 1905634638
Conc: 500.00 ng/ml