

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR062619.M
 Title : GC EXTRACTABLES
 Last Update : Thu Jun 27 06:40:13 2019
 Response Via : Initial Calibration

Calibration Files

750 =PR038914.D 1000 =PR038913.D 500 =PR038915.D
 250 =PR038916.D 50 =PR038917.D

Compound		750	1000	500	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	1.025	0.965	1.058	0.843	1.014	0.981	E6	8.57
2) SA	Decachlorobiphenyl	2.364	2.224	2.396	2.085	2.689	2.352	E6	9.59
3) L1	AR-1016-1	4.841	4.501	5.087	4.390	5.453	4.854	E4	8.94
4) L1	AR-1016-2	7.690	7.260	7.893	6.932	8.295	7.614	E4	7.00
5) L1	AR-1016-3	3.900	3.624	4.032	3.416	4.107	3.816	E4	7.59
6) L1	AR-1016-4	3.164	2.926	3.306	2.890	3.724	3.202	E4	10.56
7) L1	AR-1016-5	4.308	4.202	4.446	3.873	4.840	4.334	E4	8.16
8) L2	AR-1221-1			1.078			1.078	E4	0.00
9) L2	AR-1221-2			8.670			8.670	E3	0.00
10) L2	AR-1221-3			2.889			2.889	E4	0.00
11) L3	AR-1232-1			2.585			2.585	E4	0.00
12) L3	AR-1232-2			3.288			3.288	E4	0.00
13) L3	AR-1232-3			1.578			1.578	E4	0.00
14) L3	AR-1232-4			1.452			1.452	E4	0.00
15) L3	AR-1232-5			1.666			1.666	E4	0.00
16) L4	AR-1242-1			4.027			4.027	E4	0.00
17) L4	AR-1242-2			6.719			6.719	E4	0.00
18) L4	AR-1242-3			3.320			3.320	E4	0.00
19) L4	AR-1242-4			3.346			3.346	E4	0.00
20) L4	AR-1242-5			4.817			4.817	E4	0.00
21) L5	AR-1248-1			3.554			3.554	E4	0.00
22) L5	AR-1248-2			5.073			5.073	E4	0.00
23) L5	AR-1248-3			5.291			5.291	E4	0.00
24) L5	AR-1248-4			6.630			6.630	E4	0.00
25) L5	AR-1248-5			7.315			7.315	E4	0.00
26) L6	AR-1254-1			9.461			9.461	E4	0.00
27) L6	AR-1254-2			8.394			8.394	E4	0.00
28) L6	AR-1254-3			1.478			1.478	E5	0.00
29) L6	AR-1254-4			1.139			1.139	E5	0.00
30) L6	AR-1254-5			1.397			1.397	E5	0.00
31) L7	AR-1260-1	0.976	0.886	0.992	0.873	1.128	0.971	E5	10.53
32) L7	AR-1260-2	1.258	1.149	1.273	1.118	1.498	1.259	E5	11.87
33) L7	AR-1260-3	1.162	1.061	1.173	1.008	1.276	1.136	E5	9.21
34) L7	AR-1260-4	1.042	0.950	1.053	0.917	1.145	1.022	E5	8.84
35) L7	AR-1260-5	2.749	2.546	2.748	2.330	2.806	2.636	E5	7.49
36) L8	AR-1262-1			6.965			6.965	E4	0.00
37) L8	AR-1262-2			3.551			3.551	E5	0.00
38) L8	AR-1262-3			1.345			1.345	E5	0.00
39) L8	AR-1262-4			2.616			2.616	E5	0.00
40) L8	AR-1262-5			1.264			1.264	E5	0.00
41) L9	AR-1268-1			3.598			3.598	E5	0.00
42) L9	AR-1268-2			3.344			3.344	E5	0.00
43) L9	AR-1268-3			2.743			2.743	E5	0.00
44) L9	AR-1268-4			1.252			1.252	E5	0.00
45) L9	AR-1268-5			9.952			9.952	E5	0.00

Signal #2 Calibration Files

750 =PR038914.D 1000 =PR038913.D 500 =PR038915.D
 250 =PR038916.D 50 =PR038917.D

Compound		750	1000	500	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	3.635	3.404	3.766	3.102	3.865	3.554	E6	8.61
2) SA	Decachlorobiphenyl	7.260	6.805	7.351	6.445	8.451	7.262	E6	10.43
3) L1	AR-1016-1	1.677	1.569	1.743	1.521	2.022	1.706	E5	11.55
4) L1	AR-1016-2	2.436	2.280	2.501	2.182	2.860	2.452	E5	10.63
5) L1	AR-1016-3	1.488	1.374	1.542	1.370	1.865	1.528	E5	13.26

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR062619.M
 Title : GC EXTRACTABLES
 Last Update : Thu Jun 27 06:40:13 2019
 Response Via : Initial Calibration

Calibration Files

750 =PR038914.D 1000 =PR038913.D 500 =PR038915.D
 250 =PR038916.D 50 =PR038917.D

	Compound	750	1000	500	250	50	Avg		%RSD

6)	L1 AR-1016-4	1.223	1.138	1.250	1.095	1.327	1.207	E5	7.60
7)	L1 AR-1016-5	1.267	1.170	1.301	1.148	1.549	1.287	E5	12.44
8)	L2 AR-1221-1			3.743			3.743	E4	0.00
9)	L2 AR-1221-2			2.947			2.947	E4	0.00
10)	L2 AR-1221-3			9.466			9.466	E4	0.00
11)	L3 AR-1232-1			8.280			8.280	E4	0.00
12)	L3 AR-1232-2			4.536			4.536	E4	0.00
13)	L3 AR-1232-3			1.016			1.016	E5	0.00
14)	L3 AR-1232-4			4.990			4.990	E4	0.00
15)	L3 AR-1232-5			3.748			3.748	E4	0.00
16)	L4 AR-1242-1			1.425			1.425	E5	0.00
17)	L4 AR-1242-2			2.057			2.057	E5	0.00
18)	L4 AR-1242-3			1.264			1.264	E5	0.00
19)	L4 AR-1242-4			1.023			1.023	E5	0.00
20)	L4 AR-1242-5			1.304			1.304	E5	0.00
21)	L5 AR-1248-1			1.194			1.194	E5	0.00
22)	L5 AR-1248-2			1.640			1.640	E5	0.00
23)	L5 AR-1248-3			1.944			1.944	E5	0.00
24)	L5 AR-1248-4			2.457			2.457	E5	0.00
25)	L5 AR-1248-5			2.344			2.344	E5	0.00
26)	L6 AR-1254-1			2.039			2.039	E5	0.00
27)	L6 AR-1254-2			3.270			3.270	E5	0.00
28)	L6 AR-1254-3			3.820			3.820	E5	0.00
29)	L6 AR-1254-4			2.944			2.944	E5	0.00
30)	L6 AR-1254-5			3.259			3.259	E5	0.00
31)	L7 AR-1260-1	2.710	2.483	2.767	2.450	3.199	2.722	E5	11.03
32)	L7 AR-1260-2	3.434	3.147	3.476	3.061	3.919	3.407	E5	9.89
33)	L7 AR-1260-3	2.730	2.488	2.762	2.422	3.168	2.714	E5	10.83
34)	L7 AR-1260-4	3.281	3.007	3.311	2.877	3.724	3.240	E5	10.08
35)	L7 AR-1260-5	7.232	6.713	7.239	6.158	7.700	7.008	E5	8.42
36)	L8 AR-1262-1			4.758			4.758	E5	0.00
37)	L8 AR-1262-2			9.600			9.600	E5	0.00
38)	L8 AR-1262-3			6.387			6.387	E5	0.00
39)	L8 AR-1262-4			4.829			4.829	E5	0.00
40)	L8 AR-1262-5			3.595			3.595	E5	0.00
41)	L9 AR-1268-1			1.000			1.000	E6	0.00
42)	L9 AR-1268-2			9.421			9.421	E5	0.00
43)	L9 AR-1268-3			7.825			7.825	E5	0.00
44)	L9 AR-1268-4			3.577			3.577	E5	0.00
45)	L9 AR-1268-5			2.873			2.873	E6	0.00

(#) = Out of Range