

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR062722\
 Data File : PR054979.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Jun 2022 13:40
 Operator : AJ\MA
 Sample : AR1248ICC1000
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 27 14:38:53 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR062722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Jun 27 14:38:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.641	2.913	397.8E6	183.1E6	97.343	97.474
2) SA Decachlor...	9.534	8.142	146.6E6	147.8E6	91.716	93.486
Target Compounds						
21) L5 AR-1248-1	4.863	4.025	61904613	34614536	907.103	930.407
22) L5 AR-1248-2	5.154	4.271	72804322	44738723	891.253	922.225
23) L5 AR-1248-3	5.366	4.312	81903865	45057919	910.299	926.034
24) L5 AR-1248-4	5.798	4.486	79936761	56693682	903.442	946.353
25) L5 AR-1248-5	5.838	4.895	77098487	55761308	899.454	946.493



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR062722\
Data File : PR054979.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Jun 2022 13:40
Operator : AJ\MA
Sample : AR1248ICC1000
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 27 14:38:53 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR062722.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jun 27 14:38:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.641 min
Delta R.T.: 0.000 min
Response: 397799701
Conc: 97.34 ng/ml

Instrument :
ECD_R
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 2.913 min
Delta R.T.: 0.000 min
Response: 183139652
Conc: 97.47 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.534 min
Delta R.T.: 0.000 min
Response: 146634090
Conc: 91.72 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.142 min
Delta R.T.: 0.000 min
Response: 147763635
Conc: 93.49 ng/ml



#21 AR-1248-1


R.T.: 4.863 min
Delta R.T.: 0.000 min
Response: 61904613
Conc: 907.10 ng/ml

Instrument :
ECD_R
ClientSampleId :




#21 AR-1248-1

R.T.: 4.025 min
Delta R.T.: 0.000 min
Response: 34614536
Conc: 930.41 ng/ml




#22 AR-1248-2

R.T.: 5.154 min
Delta R.T.: 0.000 min
Response: 72804322
Conc: 891.25 ng/ml



#22 AR-1248-2


R.T.: 4.271 min
Delta R.T.: 0.000 min
Response: 44738723
Conc: 922.22 ng/ml



#23 AR-1248-3


R.T.: 5.366 min
Delta R.T.: 0.000 min
Response: 81903865
Conc: 910.30 ng/ml

Instrument :
ECD_R
ClientSampleId :




#23 AR-1248-3

R.T.: 4.312 min
Delta R.T.: 0.000 min
Response: 45057919
Conc: 926.03 ng/ml




#24 AR-1248-4

R.T.: 5.798 min
Delta R.T.: 0.000 min
Response: 79936761
Conc: 903.44 ng/ml



#24 AR-1248-4


R.T.: 4.486 min
Delta R.T.: 0.000 min
Response: 56693682
Conc: 946.35 ng/ml



#25 AR-1248-5

R.T.: 5.838 min
Delta R.T.: 0.000 min
Response: 77098487
Conc: 899.45 ng/ml

Instrument :
ECD_R
ClientSampleId :



#25 AR-1248-5

R.T.: 4.895 min
Delta R.T.: 0.000 min
Response: 55761308
Conc: 946.49 ng/ml