

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR062722\
 Data File : PR054967.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Jun 2022 10:32
 Operator : AJ\MA
 Sample : AR1660ICC1000
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR1660ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 27 12:41:58 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR062722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Jun 27 12:39:32 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.641	2.913	361.8E6	166.7E6	96.346	97.184
2) SA Decachlor...	9.534	8.143	131.0E6	131.7E6	90.448	93.718
Target Compounds						
3) L1 AR-1016-1	4.862	4.024	93776960	53638891	920.015	924.630
4) L1 AR-1016-2	4.884	4.042	133.5E6	74814436	912.364	938.149
5) L1 AR-1016-3	4.948	4.220	78187205	39858390	896.902	914.911
6) L1 AR-1016-4	5.051	4.271	64281948	30094975	900.195	899.713
7) L1 AR-1016-5	5.366	4.486	56814635	38486492	886.358	862.277
31) L7 AR-1260-1	6.578	5.571	75790490	72403751	883.697	920.066
32) L7 AR-1260-2	6.860	5.775	86930277	86573652	888.714	908.471
33) L7 AR-1260-3	7.251	5.932	59747367	82619224	904.517	915.698
34) L7 AR-1260-4	7.493	6.436	71892915	63974097	879.579	907.262
35) L7 AR-1260-5	7.832	6.701	141.2E6	157.9E6	922.224	929.345



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR062722\
Data File : PR054967.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Jun 2022 10:32
Operator : AJ\MA
Sample : AR1660ICC1000
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR1660ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 27 12:41:58 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR062722.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jun 27 12:39:32 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.641 min
Delta R.T.: 0.000 min
Response: 361793000
Conc: 96.35 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1660ICC1000




#1 Tetrachloro-m-xylene

R.T.: 2.913 min
Delta R.T.: 0.000 min
Response: 166657817
Conc: 97.18 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.534 min
Delta R.T.: 0.003 min
Response: 131011228
Conc: 90.45 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.143 min
Delta R.T.: 0.002 min
Response: 131700104
Conc: 93.72 ng/ml



#3 AR-1016-1


R.T.: 4.862 min
Delta R.T.: 0.000 min
Response: 93776960
Conc: 920.01 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1660ICC1000




#3 AR-1016-1

R.T.: 4.024 min
Delta R.T.: 0.000 min
Response: 53638891
Conc: 924.63 ng/ml




#4 AR-1016-2

R.T.: 4.884 min
Delta R.T.: 0.000 min
Response: 133454085
Conc: 912.36 ng/ml



#4 AR-1016-2


R.T.: 4.042 min
Delta R.T.: 0.000 min
Response: 74814436
Conc: 938.15 ng/ml



#5 AR-1016-3


R.T.: 4.948 min
Delta R.T.: 0.000 min
Response: 78187205
Conc: 896.90 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1660ICC1000




#5 AR-1016-3

R.T.: 4.220 min
Delta R.T.: 0.000 min
Response: 39858390
Conc: 914.91 ng/ml




#6 AR-1016-4

R.T.: 5.051 min
Delta R.T.: 0.000 min
Response: 64281948
Conc: 900.19 ng/ml



#6 AR-1016-4


R.T.: 4.271 min
Delta R.T.: 0.000 min
Response: 30094975
Conc: 899.71 ng/ml



#7 AR-1016-5


R.T.: 5.366 min
Delta R.T.: 0.000 min
Response: 56814635
Conc: 886.36 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1660ICC1000




#7 AR-1016-5

R.T.: 4.486 min
Delta R.T.: 0.000 min
Response: 38486492
Conc: 862.28 ng/ml




#31 AR-1260-1

R.T.: 6.578 min
Delta R.T.: 0.000 min
Response: 75790490
Conc: 883.70 ng/ml



#31 AR-1260-1


R.T.: 5.571 min
Delta R.T.: 0.000 min
Response: 72403751
Conc: 920.07 ng/ml



#32 AR-1260-2


R.T.: 6.860 min
Delta R.T.: 0.000 min
Response: 86930277
Conc: 888.71 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1660ICC1000




#32 AR-1260-2

R.T.: 5.775 min
Delta R.T.: 0.000 min
Response: 86573652
Conc: 908.47 ng/ml




#33 AR-1260-3

R.T.: 7.251 min
Delta R.T.: 0.002 min
Response: 59747367
Conc: 904.52 ng/ml



#33 AR-1260-3


R.T.: 5.932 min
Delta R.T.: 0.000 min
Response: 82619224
Conc: 915.70 ng/ml



#34 AR-1260-4


R.T.: 7.493 min
Delta R.T.: 0.002 min
Response: 71892915
Conc: 879.58 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1660ICC1000




#34 AR-1260-4

R.T.: 6.436 min
Delta R.T.: 0.000 min
Response: 63974097
Conc: 907.26 ng/ml



#35 AR-1260-5

R.T.: 7.832 min
Delta R.T.: 0.002 min
Response: 141160512
Conc: 922.22 ng/ml



#35 AR-1260-5

R.T.: 6.701 min
Delta R.T.: 0.000 min
Response: 157946780
Conc: 929.34 ng/ml