

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR070118\
 Data File : PR029604.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Jun 2018 10:49
 Operator : UA/SM
 Sample : J3684-03RE
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 TP-1-SRE

Manual Integrations
 APPROVED

UGO
 6/28/2018 11:53:35 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 27 11:09:49 2018
 Quant Method : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_R\Method\PR062718.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jun 23 08:41:26 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.567	3.719	322.0E6	578.7E6	13.990	12.241
2)	SA Decachlor...	10.332	8.647	305.7E6	212.0E6	10.921	10.735
Target Compounds							
31)	L7 AR-1260-1	7.318	6.241	97492042	199.1E6	79.414m	60.081m
32)	L7 AR-1260-2	7.571	6.424	157.2E6	339.9E6	100.230	80.908m
33)	L7 AR-1260-3	7.932	6.749	136.7E6	357.3E6	112.063	86.539
34)	L7 AR-1260-4	8.160	7.041	193.7E6	291.1E6	145.240	118.930
35)	L7 AR-1260-5	8.484	7.281	555.8E6	766.2E6	189.105	137.635 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR070118\
Data File : PR029604.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Jun 2018 10:49
Operator : UA/SM
Sample : J3684-03RE
Misc :
ALS Vial : 4 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampled :
TP-1-SRE


Manual Integrations
APPROVED

UGO
6/28/2018 11:53:35 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 27 11:09:49 2018
Quant Method : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_R\Method\PR062718.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jun 23 08:41:26 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.567 min
Delta R.T.: -0.006 min
Response: 321992354
Conc: 13.99 ng/ml

Instrument :
ECD_R
ClientSampleId :
TP-1-SRE


Manual Integrations
APPROVED

UGO
6/28/2018 11:53:35 AM




#1 Tetrachloro-m-xylene

R.T.: 3.719 min
Delta R.T.: 0.006 min
Response: 578715502
Conc: 12.24 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.332 min
Delta R.T.: -0.018 min
Response: 305741086
Conc: 10.92 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.647 min
Delta R.T.: -0.011 min
Response: 212001632
Conc: 10.73 ng/ml




#31 AR-1260-1

R.T.: 7.318 min
Delta R.T.: -0.011 min
Response: 97492042
Conc: 79.41 ng/ml m

Instrument :
ECD_R
ClientSampleId :
TP-1-SRE


Manual Integrations
APPROVED

UGO
6/28/2018 11:53:35 AM




#31 AR-1260-1

R.T.: 6.241 min
Delta R.T.: -0.004 min
Response: 199084312
Conc: 60.08 ng/ml m




#32 AR-1260-2

R.T.: 7.571 min
Delta R.T.: -0.013 min
Response: 157194394
Conc: 100.23 ng/ml



#32 AR-1260-2

R.T.: 6.424 min
Delta R.T.: -0.006 min
Response: 339933907
Conc: 80.91 ng/ml m




#33 AR-1260-3

R.T.: 7.932 min
Delta R.T.: -0.011 min
Response: 136655192
Conc: 112.06 ng/ml

Instrument :
ECD_R
ClientSampleId :
TP-1-SRE


Manual Integrations
APPROVED

UGO
6/28/2018 11:53:35 AM




#33 AR-1260-3

R.T.: 6.749 min
Delta R.T.: -0.006 min
Response: 357284718
Conc: 86.54 ng/ml




#34 AR-1260-4

R.T.: 8.160 min
Delta R.T.: -0.011 min
Response: 193712967
Conc: 145.24 ng/ml



#34 AR-1260-4

R.T.: 7.041 min
Delta R.T.: -0.007 min
Response: 291146294
Conc: 118.93 ng/ml




#35 AR-1260-5

R.T.: 8.484 min
Delta R.T.: -0.012 min
Response: 555798184
Conc: 189.10 ng/ml

Instrument :
ECD_R
ClientSampleId :
TP-1-SRE

Manual Integrations
APPROVED

UGO
6/28/2018 11:53:35 AM



#35 AR-1260-5

R.T.: 7.281 min
Delta R.T.: -0.007 min
Response: 766218570
Conc: 137.63 ng/ml